

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN112724

Review By	yogesh	Review On	11/29/2024 12:48:33 AM
Supervise By	mohammad	Supervise On	12/3/2024 12:32:57 AM
SubDirectory	BN112724	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn112724
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657		
CCC	SP6661		
Internal Standard/PEM	SP6527,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN035349.D	27 Nov 2024 14:55	RC/JU	Ok
2	SSTDICC0.1	BN035350.D	27 Nov 2024 15:34	RC/JU	Ok,M
3	SSTDICC0.2	BN035351.D	27 Nov 2024 16:10	RC/JU	Ok,M
4	SSTDICCC0.4	BN035352.D	27 Nov 2024 16:46	RC/JU	Ok,M
5	SSTDICC0.8	BN035353.D	27 Nov 2024 17:21	RC/JU	Ok
6	SSTDICC1.6	BN035354.D	27 Nov 2024 17:57	RC/JU	Ok
7	SSTDICC3.2	BN035355.D	27 Nov 2024 18:33	RC/JU	Ok
8	SSTDICC5.0	BN035356.D	27 Nov 2024 19:09	RC/JU	Ok
9	SSTDICV0.4	BN035357.D	27 Nov 2024 20:21	RC/JU	Ok,M
10	PB165224BL	BN035358.D	27 Nov 2024 20:57	RC/JU	Ok,M
11	PB165266BL	BN035359.D	27 Nov 2024 21:32	RC/JU	Ok,M
12	PB165224BS	BN035360.D	27 Nov 2024 22:08	RC/JU	Ok,M
13	PB165224BSD	BN035361.D	27 Nov 2024 22:44	RC/JU	Ok,M
14	P4969-02	BN035362.D	27 Nov 2024 23:20	RC/JU	Ok,M
15	P4969-04	BN035363.D	27 Nov 2024 23:56	RC/JU	Ok,M
16	P5003-02	BN035364.D	28 Nov 2024 00:32	RC/JU	Ok,M
17	P5003-03	BN035365.D	28 Nov 2024 01:08	RC/JU	Ok,M
18	SSTDCCC0.4	BN035366.D	28 Nov 2024 02:20	RC/JU	Ok,M
19	DFTPP	BN035367.D	28 Nov 2024 03:35	RC/JU	Ok
20	SSTDCCC0.4	BN035368.D	28 Nov 2024 04:51	RC/JU	Ok,M
21	PB165198BL	BN035369.D	28 Nov 2024 05:27	RC/JU	Ok,M

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN112724

Review By	yogesh	Review On	11/29/2024 12:48:33 AM
Supervise By	mohammad	Supervise On	12/3/2024 12:32:57 AM
SubDirectory	BN112724	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn112724
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657		
CCC	SP6661		
Internal Standard/PEM	SP6527,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB165198BS	BN035370.D	28 Nov 2024 06:03	RC/JU	Ok,M
23	PB165198BSD	BN035371.D	28 Nov 2024 06:39	RC/JU	Ok,M
24	PB165266BS	BN035372.D	28 Nov 2024 07:15	RC/JU	Ok,M
25	PB165266BSD	BN035373.D	28 Nov 2024 07:50	RC/JU	Ok,M
26	P4959-01	BN035374.D	28 Nov 2024 08:26	RC/JU	Ok,M
27	SSTDCCC0.4	BN035375.D	28 Nov 2024 09:02	RC/JU	Ok,M

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN121924

Review By	yogesh	Review On	12/21/2024 1:06:44 AM		
Supervise By	mohammad	Supervise On	12/24/2024 11:31:00 PM		
SubDirectory	BN121924	HP Acquire Method	BNA_N	HP Processing Method	bn112724
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657			
CCC		SP6661			
Internal Standard/PEM		SP6527,1ul/100ul sample			
ICV/I.BLK		SP6684			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN035711.D	19 Dec 2024 18:44	RC/JU	Ok
2	SSTDCCC0.4	BN035712.D	19 Dec 2024 19:23	RC/JU	Ok,M
3	PB165596BL	BN035713.D	19 Dec 2024 20:35	RC/JU	Ok,M
4	P5261-01	BN035714.D	19 Dec 2024 21:11	RC/JU	Ok,M
5	P5261-02	BN035715.D	19 Dec 2024 21:47	RC/JU	Ok,M
6	P5261-03	BN035716.D	19 Dec 2024 22:23	RC/JU	Ok,M
7	P5261-04	BN035717.D	19 Dec 2024 22:58	RC/JU	Dilution
8	P5261-05	BN035718.D	19 Dec 2024 23:34	RC/JU	Ok,M
9	P5261-06	BN035719.D	20 Dec 2024 00:10	RC/JU	Ok,M
10	P5261-07	BN035720.D	20 Dec 2024 00:46	RC/JU	Ok,M
11	P5261-08	BN035721.D	20 Dec 2024 01:22	RC/JU	Ok,M
12	P5261-09	BN035722.D	20 Dec 2024 01:58	RC/JU	Dilution
13	P5261-10	BN035723.D	20 Dec 2024 02:34	RC/JU	Ok,M
14	P5261-11	BN035724.D	20 Dec 2024 03:10	RC/JU	Ok,M
15	P5261-16	BN035725.D	20 Dec 2024 03:46	RC/JU	Ok,M
16	P5261-15	BN035726.D	20 Dec 2024 04:22	RC/JU	Dilution
17	P5261-17	BN035727.D	20 Dec 2024 04:58	RC/JU	Ok,M
18	PB165596BS	BN035728.D	20 Dec 2024 05:34	RC/JU	Ok,M
19	SSTDCCC0.4	BN035729.D	20 Dec 2024 06:10	RC/JU	Ok,M
20	DFTPP	BN035730.D	20 Dec 2024 07:25	RC/JU	Ok
21	SSTDCCC0.4	BN035731.D	20 Dec 2024 08:04	RC/JU	Ok,M

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN121924

Review By	yogesh	Review On	12/21/2024 1:06:44 AM		
Supervise By	mohammad	Supervise On	12/24/2024 11:31:00 PM		
SubDirectory	BN121924	HP Acquire Method	BNA_N	HP Processing Method	bn112724
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657			
CCC		SP6661			
Internal Standard/PEM		SP6527,1ul/100ul sample			
ICV/I.BLK		SP6684			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	PB165667BL	BN035732.D	20 Dec 2024 08:40	RC/JU	Ok,M
23	P5280-03	BN035733.D	20 Dec 2024 09:16	RC/JU	Not Ok
24	P5261-12	BN035734.D	20 Dec 2024 10:29	RC/JU	Ok,M
25	P5280-04MS	BN035735.D	20 Dec 2024 11:05	RC/JU	Ok,M
26	P5280-05MSD	BN035736.D	20 Dec 2024 11:41	RC/JU	Ok,M
27	P5261-13MS	BN035737.D	20 Dec 2024 12:17	RC/JU	Ok,M
28	P5261-14MSD	BN035738.D	20 Dec 2024 12:53	RC/JU	Ok,M
29	P5261-18	BN035739.D	20 Dec 2024 13:29	RC/JU	Ok,M
30	P5261-19	BN035740.D	20 Dec 2024 14:05	RC/JU	Dilution
31	P5261-20	BN035741.D	20 Dec 2024 14:41	RC/JU	Ok,M
32	P5281-01	BN035742.D	20 Dec 2024 15:18	RC/JU	Ok,M
33	P5317-01	BN035743.D	20 Dec 2024 15:54	RC/JU	Ok,M
34	P5317-02	BN035744.D	20 Dec 2024 16:30	RC/JU	Ok,M
35	P5317-03	BN035745.D	20 Dec 2024 17:06	RC/JU	Ok,M
36	P5317-04	BN035746.D	20 Dec 2024 17:42	RC/JU	Ok,M
37	SSTDCCC0.4	BN035747.D	20 Dec 2024 18:18	RC/JU	Ok,M
38	DFTPP	BN035748.D	20 Dec 2024 18:54	RC/JU	Ok
39	SSTDCCC0.4	BN035749.D	20 Dec 2024 19:33	RC/JU	Ok,M
40	PB165723BL	BN035750.D	20 Dec 2024 20:09	RC/JU	Ok,M
41	P5317-05	BN035751.D	20 Dec 2024 20:45	RC/JU	Ok,M
42	P5261-09DL	BN035752.D	20 Dec 2024 21:21	RC/JU	Ok,M
43	P5261-15DL	BN035753.D	20 Dec 2024 21:57	RC/JU	Ok,M
44	P5261-04DL	BN035754.D	20 Dec 2024 22:33	RC/JU	Ok,M

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN121924

Review By	yogesh	Review On	12/21/2024 1:06:44 AM		
Supervise By	mohammad	Supervise On	12/24/2024 11:31:00 PM		
SubDirectory	BN121924	HP Acquire Method	BNA_N	HP Processing Method	bn112724
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657			
CCC		SP6661			
Internal Standard/PEM		SP6527,1ul/100ul sample			
ICV/I.BLK		SP6684			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

45	P5317-09	BN035755.D	20 Dec 2024 23:09	RC/JU	Ok,M
46	P5317-10	BN035756.D	20 Dec 2024 23:45	RC/JU	Ok,M
47	P5317-11	BN035757.D	21 Dec 2024 00:21	RC/JU	Ok,M
48	P5317-12	BN035758.D	21 Dec 2024 00:57	RC/JU	Ok,M
49	P5317-08	BN035759.D	21 Dec 2024 01:34	RC/JU	Ok,M
50	P5317-13	BN035760.D	21 Dec 2024 02:10	RC/JU	Ok,M
51	P5317-14	BN035761.D	21 Dec 2024 02:46	RC/JU	Ok,M
52	P5284-02	BN035762.D	21 Dec 2024 03:22	RC/JU	Ok,M
53	P5284-03	BN035763.D	21 Dec 2024 03:58	RC/JU	Ok,M
54	PB165724BS	BN035764.D	21 Dec 2024 04:34	RC/JU	Ok,M
55	SSTDCCC0.4	BN035765.D	21 Dec 2024 05:10	RC/JU	Ok,M
56	DFTPP	BN035766.D	21 Dec 2024 05:46	RC/JU	Ok
57	SSTDCCC0.4	BN035767.D	21 Dec 2024 06:25	RC/JU	Ok,M
58	PB165596BL	BN035768.D	21 Dec 2024 07:01	RC/JU	Not Ok
59	P5281-04	BN035769.D	21 Dec 2024 07:37	RC/JU	Not Ok
60	P5317-15	BN035770.D	21 Dec 2024 08:13	RC/JU	Not Ok
61	P5317-16	BN035771.D	21 Dec 2024 08:49	RC/JU	Not Ok
62	P5317-17	BN035772.D	21 Dec 2024 09:25	RC/JU	Not Ok
63	P5317-18	BN035773.D	21 Dec 2024 10:01	RC/JU	Not Ok
64	P5317-19MS	BN035774.D	21 Dec 2024 10:37	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN122624

Review By	yogesh	Review On	12/27/2024 3:53:55 AM
Supervise By	mohammad	Supervise On	12/27/2024 4:13:15 AM
SubDirectory	BN122624	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn112724
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657		
CCC	SP6661		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN035844.D	26 Dec 2024 09:19	RC/JU	Ok
2	SSTDCCC0.4	BN035845.D	26 Dec 2024 09:58	RC/JU	Ok,M
3	PB165828BL	BN035846.D	26 Dec 2024 10:35	RC/JU	Ok,M
4	P5376-15	BN035847.D	26 Dec 2024 11:20	RC/JU	Ok,M
5	P5376-19	BN035848.D	26 Dec 2024 11:56	RC/JU	Ok,M
6	P5376-01	BN035849.D	26 Dec 2024 12:31	RC/JU	Ok,M
7	P5374-03	BN035850.D	26 Dec 2024 13:07	RC/JU	Ok,M
8	PB165723BS	BN035851.D	26 Dec 2024 13:43	RC/JU	Ok,M
9	PB165828BS	BN035852.D	26 Dec 2024 14:19	RC/JU	Ok,M
10	PB165724BSD	BN035853.D	26 Dec 2024 14:54	RC/JU	Ok,M
11	P5376-03DL	BN035854.D	26 Dec 2024 15:30	RC/JU	Ok,M
12	P5317-21DL	BN035855.D	26 Dec 2024 16:06	RC/JU	Ok,M
13	P5281-07DL	BN035856.D	26 Dec 2024 16:42	RC/JU	Ok,M
14	P5282-01DL	BN035857.D	26 Dec 2024 17:18	RC/JU	Ok,M
15	P5281-06DL	BN035858.D	26 Dec 2024 17:53	RC/JU	Ok,M
16	P5284-01DL	BN035859.D	26 Dec 2024 18:29	RC/JU	Ok,M
17	P5261-19DL	BN035860.D	26 Dec 2024 19:05	RC/JU	Ok,M
18	SSTDCCC0.4	BN035861.D	26 Dec 2024 19:41	RC/JU	Ok,M

M : Manual Integration