

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG NO.: P5261

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/19/2024

Lab File ID: BN035712.D Time Analyzed: 19:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3021	7.272	8329	10.02	5302	13.94
UPPER LIMIT	6042	7.772	16658	10.52	10604	14.435
LOWER LIMIT	1510.5	6.772	4164.5	9.52	2651	13.435
EPA SAMPLE NO.						
01 RE122D1-20241209	3160	7.27	8794	10.02	6070	13.94
02 RE122D2-20241209	3267	7.27	8893	10.02	5822	13.94
03 RE122D3-20241209	3243	7.27	9120	10.02	6257	13.94
04 RE103D1-20241209	3204	7.27	8678	10.02	5798	13.94
05 RE103D2-20241209	3198	7.27	8727	10.02	5748	13.94
06 RE103D3-20241209	2937	7.27	7894	10.02	4865	13.94
07 RE134D1-20241209	2993	7.27	8239	10.02	5412	13.94
08 DUP-01-20241209	3370	7.27	8947	10.02	5786	13.94
09 RE134D3-20241210	3093	7.27	8301	10.02	5668	13.94
10 RE134D4-20241210	2978	7.27	8067	10.02	5314	13.94
11 GM3-RW3-MW1-20241210	3049	7.27	8135	10.02	5075	13.94
12 RE125D2-20241210	2948	7.27	7525	10.02	4789	13.94
13 RE125D3-20241210	2841	7.27	7711	10.02	5101	13.94
14 RE120D1-20241210	2945	7.27	7549	10.02	4981	13.94
15 PB165596BL	3228	7.27	8561	10.02	5768	13.94
16 PB165596BS	2951	7.27	7713	10.02	4640	13.94

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG NO.: P5261

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/19/2024

Lab File ID: BN035712.D Time Analyzed: 19:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	3021	7.272	8329	10.02	5302	13.94
UPPER LIMIT	6042	7.772	16658	10.52	10604	14.435
LOWER LIMIT	1510.5	6.772	4164.5	9.52	2651	13.435
EPA SAMPLE NO.						

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.