

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: P5261 SDG NO.: P5261

Lab File ID: BN035711.D

DFTPP Injection Date: 12/19/2024

Instrument ID: BNA_N

DFTPP Injection Time: 18:44

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.5
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	33
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	40.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	29.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	13.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	15.4 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035712.D	12/19/2024	19:23
PB165596BL	PB165596BL	BN035713.D	12/19/2024	20:35
RE122D1-20241209	P5261-01	BN035714.D	12/19/2024	21:11
RE122D2-20241209	P5261-02	BN035715.D	12/19/2024	21:47
RE122D3-20241209	P5261-03	BN035716.D	12/19/2024	22:23
RE103D1-20241209	P5261-04	BN035717.D	12/19/2024	22:58
RE103D2-20241209	P5261-05	BN035718.D	12/19/2024	23:34
RE103D3-20241209	P5261-06	BN035719.D	12/20/2024	00:10
RE134D1-20241209	P5261-07	BN035720.D	12/20/2024	00:46
DUP-01-20241209	P5261-08	BN035721.D	12/20/2024	01:22
RE134D3-20241210	P5261-09	BN035722.D	12/20/2024	01:58
RE134D4-20241210	P5261-10	BN035723.D	12/20/2024	02:34
GM3-RW3-MW1-20241210	P5261-11	BN035724.D	12/20/2024	03:10
RE125D3-20241210	P5261-16	BN035725.D	12/20/2024	03:46
RE125D2-20241210	P5261-15	BN035726.D	12/20/2024	04:22
RE120D1-20241210	P5261-17	BN035727.D	12/20/2024	04:58
PB165596BS	PB165596BS	BN035728.D	12/20/2024	05:34
SSTDCCC0.4EC	SSTDCCC0.4	BN035729.D	12/20/2024	06:10