

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG No.: P5261
 Instrument ID: MSVOA_X Calibration Date(s): 12/11/2024 12/11/2024
 Heated Purge: (Y/N) N Calibration Time(s): 10:41 13:00
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX044219.D		RRF005 = VX044220.D		RRF020 = VX044221.D		RRF050 = VX044222.D		RRF100 = VX044223.D		RRF150 = VX044224.D	
COMPOUND		RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF		% RSD			
Dichlorodifluoromethane		0.877	0.749	0.721	0.788	0.745	0.759	0.773		7.2			
Chloromethane		0.743	0.747	0.735	0.756	0.715	0.740	0.739		1.8			
Vinyl Chloride		0.805	0.790	0.782	0.786	0.722	0.737	0.770		4.3			
Bromomethane			0.590	0.513	0.551	0.529	0.548	0.546		5.3			
Chloroethane		0.464	0.520	0.515	0.532	0.422	0.426	0.480		10.2			
Trichlorofluoromethane		1.582	1.519	1.514	1.541	1.478	1.414	1.508		3.8			
1,1,2-Trichlorotrifluoroethane		0.633	0.628	0.592	0.615	0.573	0.620	0.610		3.8			
1,1-Dichloroethene		0.512	0.560	0.544	0.596	0.554	0.596	0.560		5.7			
Acetone		0.359	0.353	0.331	0.381	0.348	0.372	0.357		5			
Carbon Disulfide		0.841	0.872	0.937	1.150	1.230	1.362	1.065		20			
Methyl tert-butyl Ether		2.054	2.188	2.127	2.305	2.170	2.279	2.187		4.3			
Methyl Acetate		0.965	0.889	0.883	0.953	0.888	0.951	0.922		4.2			
Methylene Chloride		0.679	0.678	0.666	0.689	0.648	0.668	0.671		2.1			
trans-1,2-Dichloroethene		0.600	0.556	0.577	0.619	0.595	0.626	0.596		4.4			
1,1-Dichloroethane		1.071	1.173	1.154	1.244	1.171	1.218	1.172		5.1			
Cyclohexane			0.931	0.963	0.977	0.922	0.938	0.946		2.4			
2-Butanone		0.449	0.491	0.510	0.548	0.509	0.541	0.508		7.1			
Carbon Tetrachloride		0.484	0.443	0.432	0.501	0.502	0.529	0.482		7.8			
cis-1,2-Dichloroethene		0.809	0.737	0.728	0.772	0.731	0.762	0.756		4.1			
Bromochloromethane		0.578	0.577	0.576	0.592	0.556	0.568	0.575		2.1			
Chloroform		1.389	1.312	1.272	1.343	1.272	1.310	1.316		3.4			
1,1,1-Trichloroethane		0.993	1.019	1.077	1.176	1.119	1.173	1.093		7.1			
Methylcyclohexane		0.514	0.503	0.527	0.555	0.550	0.562	0.535		4.5			
Benzene		1.353	1.353	1.332	1.410	1.353	1.353	1.359		1.9			
1,2-Dichloroethane		0.554	0.557	0.545	0.584	0.556	0.564	0.560		2.4			
Trichloroethene		0.356	0.325	0.325	0.349	0.337	0.342	0.339		3.8			
1,2-Dichloropropane		0.368	0.345	0.330	0.357	0.341	0.342	0.347		3.8			
Bromodichloromethane		0.380	0.411	0.436	0.508	0.518	0.538	0.465		13.9			
4-Methyl-2-Pentanone		0.506	0.551	0.547	0.590	0.561	0.582	0.556		5.4			
Toluene		0.908	0.830	0.816	0.877	0.841	0.849	0.853		4			

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG No.: P5261
Instrument ID: MSVOA_X Calibration Date(s): 12/11/2024 12/11/2024
Heated Purge: (Y/N) N Calibration Time(s): 10:41 13:00
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX044219.D		RRF005 = VX044220.D		RRF020 = VX044221.D			
		RRF050 = VX044222.D		RRF100 = VX044223.D		RRF150 = VX044224.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.338	0.357	0.439	0.514	0.526	0.558	0.455	20.3	
cis-1,3-Dichloropropene	0.404	0.441	0.488	0.567	0.558	0.583	0.507	14.6	
1,1,2-Trichloroethane	0.337	0.332	0.335	0.358	0.334	0.340	0.339	2.8	
2-Hexanone	0.323	0.386	0.408	0.439	0.420	0.438	0.403	10.8	
Dibromochloromethane	0.254	0.248	0.306	0.370	0.378	0.398	0.326	20.2	
1,2-Dibromoethane	0.303	0.322	0.344	0.369	0.364	0.370	0.345	8.1	
Tetrachloroethene	0.359	0.354	0.317	0.321	0.312	0.327	0.332	6	
Chlorobenzene	1.130	1.068	1.039	1.088	1.036	1.067	1.071	3.3	
Ethyl Benzene	1.853	1.806	1.816	1.927	1.831	1.875	1.851	2.4	
m/p-Xylenes	0.653	0.660	0.665	0.707	0.683	0.707	0.679	3.5	
o-Xylene	0.686	0.683	0.685	0.703	0.671	0.697	0.687	1.6	
Styrene	1.007	1.019	1.099	1.162	1.148	1.177	1.102	6.7	
Bromoform	0.137	0.173	0.194	0.247	0.269	0.297	0.219	28	
Isopropylbenzene	3.665	3.976	3.920	3.912	3.747	3.747	3.828	3.2	
1,1,2,2-Tetrachloroethane	1.410	1.350	1.332	1.336	1.256	1.293	1.329	3.9	
1,3-Dichlorobenzene	1.519	1.634	1.640	1.683	1.625	1.628	1.621	3.3	
1,4-Dichlorobenzene	1.811	1.640	1.632	1.672	1.621	1.623	1.667	4.4	
1,2-Dichlorobenzene	1.745	1.666	1.668	1.726	1.654	1.643	1.684	2.5	
1,2-Dibromo-3-Chloropropane	0.222	0.226	0.241	0.273	0.279	0.306	0.258	13	
1,2,4-Trichlorobenzene	0.770	0.827	0.891	0.978	1.013	1.058	0.923	12.2	
1,2,3-Trichlorobenzene	0.854	0.885	0.984	1.019	1.055	1.077	0.979	9.3	
1,2-Dichloroethane-d4		0.934	0.778	0.921	0.862	0.888	0.877	7.1	
Dibromofluoromethane		0.353	0.298	0.367	0.360	0.354	0.346	8	
Toluene-d8		1.148	1.010	1.251	1.229	1.203	1.168	8.3	
4-Bromofluorobenzene		0.390	0.351	0.442	0.432	0.426	0.408	9.2	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.