

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG No.: P5261

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 11:57

Lab File ID: VX044375.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.800		3.49	20
Chloromethane	0.739	0.720	0.1	-2.57	20
Vinyl Chloride	0.770	0.743		-3.51	20
Bromomethane	0.546	0.475		-13	20
Chloroethane	0.480	0.494		2.92	20
Trichlorofluoromethane	1.508	1.275		-15.45	20
1,1,2-Trichlorotrifluoroethane	0.610	0.613		0.49	20
1,1-Dichloroethene	0.560	0.566		1.07	20
Acetone	0.357	0.298		-16.53	20
Carbon Disulfide	1.065	1.209		13.52	20
Methyl tert-butyl Ether	2.187	2.034		-7	20
Methyl Acetate	0.922	0.830		-9.98	20
Methylene Chloride	0.671	0.643		-4.17	20
trans-1,2-Dichloroethene	0.596	0.594		-0.34	20
1,1-Dichloroethane	1.172	1.124	0.1	-4.1	20
Cyclohexane	0.946	0.956		1.06	20
2-Butanone	0.508	0.450		-11.42	20
Carbon Tetrachloride	0.482	0.517		7.26	20
cis-1,2-Dichloroethene	0.756	0.723		-4.36	20
Bromochloromethane	0.575	0.531		-7.65	20
Chloroform	1.316	1.196		-9.12	20
1,1,1-Trichloroethane	1.093	1.072		-1.92	20
Methylcyclohexane	0.535	0.624		16.64	20
Benzene	1.359	1.428		5.08	20
1,2-Dichloroethane	0.560	0.555		-0.89	20
Trichloroethene	0.339	0.364		7.38	20
1,2-Dichloropropane	0.347	0.356		2.59	20
Bromodichloromethane	0.465	0.508		9.25	20
4-Methyl-2-Pentanone	0.556	0.546		-1.8	20
Toluene	0.853	0.900		5.51	20
t-1,3-Dichloropropene	0.455	0.521		14.51	20
cis-1,3-Dichloropropene	0.507	0.571		12.62	20
1,1,2-Trichloroethane	0.339	0.351		3.54	20
2-Hexanone	0.403	0.402		-0.25	20
Dibromochloromethane	0.326	0.365		11.96	20
1,2-Dibromoethane	0.345	0.365		5.8	20
Tetrachloroethene	0.332	0.358		7.83	20
Chlorobenzene	1.071	1.111	0.3	3.73	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG No.: P5261

Instrument ID: MSVOA_X Calibration Date/Time: 12/18/2024 11:57

Lab File ID: VX044375.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.955		5.62	20
m/p-Xylenes	0.679	0.737		8.54	20
o-Xylene	0.687	0.731		6.41	20
Styrene	1.102	1.210		9.8	20
Bromoform	0.219	0.261	0.1	19.18	20
Isopropylbenzene	3.828	4.011		4.78	20
1,1,2,2-Tetrachloroethane	1.329	1.294	0.3	-2.63	20
1,3-Dichlorobenzene	1.621	1.757		8.39	20
1,4-Dichlorobenzene	1.667	1.734		4.02	20
1,2-Dichlorobenzene	1.684	1.744		3.56	20
1,2-Dibromo-3-Chloropropane	0.258	0.244		-5.43	20
1,2,4-Trichlorobenzene	0.923	1.075		16.47	20
1,2,3-Trichlorobenzene	0.979	1.107		13.07	20
1,2-Dichloroethane-d4	0.877	0.742		-15.39	20
Dibromofluoromethane	0.346	0.340		-1.73	20
Toluene-d8	1.168	1.164		-0.34	20
4-Bromofluorobenzene	0.408	0.413		1.23	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.