

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG No.: P5261

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 09:16

Lab File ID: VX044438.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.854		10.48	20
Chloromethane	0.739	0.782	0.1	5.82	20
Vinyl Chloride	0.770	0.798		3.64	20
Bromomethane	0.546	0.532		-2.56	20
Chloroethane	0.480	0.473		-1.46	20
Trichlorofluoromethane	1.508	1.482		-1.72	20
1,1,2-Trichlorotrifluoroethane	0.610	0.676		10.82	20
1,1-Dichloroethene	0.560	0.624		11.43	20
Acetone	0.357	0.291		-18.49	20
Carbon Disulfide	1.065	1.106		3.85	20
Methyl tert-butyl Ether	2.187	2.093		-4.3	20
Methyl Acetate	0.922	0.896		-2.82	20
Methylene Chloride	0.671	0.689		2.68	20
trans-1,2-Dichloroethene	0.596	0.625		4.87	20
1,1-Dichloroethane	1.172	1.203	0.1	2.64	20
Cyclohexane	0.946	1.000		5.71	20
2-Butanone	0.508	0.457		-10.04	20
Carbon Tetrachloride	0.482	0.533		10.58	20
cis-1,2-Dichloroethene	0.756	0.755		-0.13	20
Bromochloromethane	0.575	0.553		-3.83	20
Chloroform	1.316	1.265		-3.88	20
1,1,1-Trichloroethane	1.093	1.099		0.55	20
Methylcyclohexane	0.535	0.648		21.12	20
Benzene	1.359	1.529		12.51	20
1,2-Dichloroethane	0.560	0.586		4.64	20
Trichloroethene	0.339	0.388		14.45	20
1,2-Dichloropropane	0.347	0.369		6.34	20
Bromodichloromethane	0.465	0.525		12.9	20
4-Methyl-2-Pentanone	0.556	0.573		3.06	20
Toluene	0.853	0.924		8.32	20
t-1,3-Dichloropropene	0.455	0.517		13.63	20
cis-1,3-Dichloropropene	0.507	0.577		13.81	20
1,1,2-Trichloroethane	0.339	0.374		10.32	20
2-Hexanone	0.403	0.414		2.73	20
Dibromochloromethane	0.326	0.381		16.87	20
1,2-Dibromoethane	0.345	0.385		11.59	20
Tetrachloroethene	0.332	0.378		13.85	20
Chlorobenzene	1.071	1.187	0.3	10.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG No.: P5261

Instrument ID: MSVOA_X Calibration Date/Time: 12/20/2024 09:16

Lab File ID: VX044438.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	2.034		9.89	20
m/p-Xylenes	0.679	0.780		14.88	20
o-Xylene	0.687	0.748		8.88	20
Styrene	1.102	1.275		15.7	20
Bromoform	0.219	0.266	0.1	21.46	20
Isopropylbenzene	3.828	4.095		6.97	20
1,1,2,2-Tetrachloroethane	1.329	1.387	0.3	4.36	20
1,3-Dichlorobenzene	1.621	1.811		11.72	20
1,4-Dichlorobenzene	1.667	1.781		6.84	20
1,2-Dichlorobenzene	1.684	1.820		8.08	20
1,2-Dibromo-3-Chloropropane	0.258	0.253		-1.94	20
1,2,4-Trichlorobenzene	0.923	1.098		18.96	20
1,2,3-Trichlorobenzene	0.979	1.140		16.44	20
1,2-Dichloroethane-d4	0.877	0.836		-4.68	20
Dibromofluoromethane	0.346	0.392		13.3	20
Toluene-d8	1.168	1.318		12.84	20
4-Bromofluorobenzene	0.408	0.454		11.27	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.