

Hit Summary Sheet
SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	RE122D1-20241209								
P5261-01	RE122D1-20241209	Water	1,1,2-Trichlorotrifluoroethane	1.90		0.25	0.50	1.00	ug/L
P5261-01	RE122D1-20241209	Water	1,1-Dichloroethene	0.43	J	0.26	0.75	1.00	ug/L
P5261-01	RE122D1-20241209	Water	trans-1,2-Dichloroethene	0.40	J	0.25	0.50	1.00	ug/L
P5261-01	RE122D1-20241209	Water	cis-1,2-Dichloroethene	1.50		0.25	0.75	1.00	ug/L
P5261-01	RE122D1-20241209	Water	Trichloroethene	260	E	0.32	0.75	1.00	ug/L
P5261-01	RE122D1-20241209	Water	Tetrachloroethene	1.60		0.25	0.50	1.00	ug/L
			Total Voc :	266					
			Total Concentration:	266					
Client ID:	RE122D1-20241209DL								
P5261-01DL	RE122D1-20241209	Water	Trichloroethene	290	D	3.20	7.50	10.0	ug/L
			Total Voc :	290					
			Total Concentration:	290					
Client ID:	RE122D2-20241209								
P5261-02	RE122D2-20241209	Water	Trichloroethene	3500		32.0	75.0	100	ug/L
			Total Voc :	3500					
			Total Concentration:	3500					
Client ID:	RE122D3-20241209								
P5261-03	RE122D3-20241209	Water	Trichloroethene	3.20		0.32	0.75	1.00	ug/L
P5261-03	RE122D3-20241209	Water	m/p-Xylenes	0.42	J	0.31	1.00	2.00	ug/L
			Total Voc :	3.62					
			Total Concentration:	3.62					
Client ID:	RE103D1-20241209								
P5261-04	RE103D1-20241209	Water	1,1,2-Trichlorotrifluoroethane	8.80		0.25	0.50	1.00	ug/L
P5261-04	RE103D1-20241209	Water	1,1-Dichloroethene	1.40		0.26	0.75	1.00	ug/L
P5261-04	RE103D1-20241209	Water	cis-1,2-Dichloroethene	1.50		0.25	0.75	1.00	ug/L
P5261-04	RE103D1-20241209	Water	Chloroform	0.40	J	0.26	0.50	1.00	ug/L
P5261-04	RE103D1-20241209	Water	Trichloroethene	600	E	0.32	0.75	1.00	ug/L
P5261-04	RE103D1-20241209	Water	Tetrachloroethene	6.50		0.25	0.50	1.00	ug/L
			Total Voc :	619					
			Total Concentration:	619					
Client ID:	RE103D1-20241209DL								
P5261-04DL	RE103D1-20241209	Water	1,1,2-Trichlorotrifluoroethane	31.6	D	5.00	10.0	20.0	ug/L
P5261-04DL	RE103D1-20241209	Water	Trichloroethene	560	D	6.40	15.0	20.0	ug/L
P5261-04DL	RE103D1-20241209	Water	Tetrachloroethene	15.6	JD	5.00	10.0	20.0	ug/L
			Total Voc :	607					
			Total Concentration:	607					
Client ID:	RE103D2-20241209								
P5261-05	RE103D2-20241209	Water	1,1,2-Trichlorotrifluoroethane	3.00		0.25	0.50	1.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
P5261-05	RE103D2-20241209	Water	Acetone	1.40	J	1.40	3.80	5.00	ug/L
P5261-05	RE103D2-20241209	Water	Trichloroethene	320	E	0.32	0.75	1.00	ug/L
P5261-05	RE103D2-20241209	Water	Tetrachloroethene	0.88	J	0.25	0.50	1.00	ug/L
			Total Voc :	325					
			Total Concentration:	325					
Client ID:	RE103D2-20241209DL								
P5261-05DL	RE103D2-20241209	Water	Trichloroethene	340	D	3.20	7.50	10.0	ug/L
			Total Voc :	340					
			Total Concentration:	340					
Client ID:	RE103D3-20241209								
P5261-06	RE103D3-20241209	Water	Chloromethane	0.65	J	0.35	0.50	1.00	ug/L
P5261-06	RE103D3-20241209	Water	1,1,2-Trichlorotrifluoroethane	1.20		0.25	0.50	1.00	ug/L
P5261-06	RE103D3-20241209	Water	Acetone	1.50	J	1.40	3.80	5.00	ug/L
P5261-06	RE103D3-20241209	Water	Trichloroethene	100		0.32	0.75	1.00	ug/L
P5261-06	RE103D3-20241209	Water	Tetrachloroethene	0.36	J	0.25	0.50	1.00	ug/L
			Total Voc :	104					
			Total Concentration:	104					
Client ID:	RE134D1-20241209								
P5261-07	RE134D1-20241209	Water	Acetone	1.50	J	1.40	3.80	5.00	ug/L
P5261-07	RE134D1-20241209	Water	Chloroform	0.43	J	0.26	0.50	1.00	ug/L
P5261-07	RE134D1-20241209	Water	Trichloroethene	4.80		0.32	0.75	1.00	ug/L
P5261-07	RE134D1-20241209	Water	Tetrachloroethene	0.48	J	0.25	0.50	1.00	ug/L
			Total Voc :	7.21					
			Total Concentration:	7.21					
Client ID:	DUP-01-20241209								
P5261-08	DUP-01-20241209	Water	Methylene Chloride	0.32	J	0.32	0.50	1.00	ug/L
P5261-08	DUP-01-20241209	Water	Chloroform	0.44	J	0.26	0.50	1.00	ug/L
P5261-08	DUP-01-20241209	Water	Trichloroethene	3.20		0.32	0.75	1.00	ug/L
P5261-08	DUP-01-20241209	Water	Tetrachloroethene	0.48	J	0.25	0.50	1.00	ug/L
			Total Voc :	4.44					
			Total Concentration:	4.44					
Client ID:	RE134D3-20241210								
P5261-09	RE134D3-20241210	Water	1,1,2-Trichlorotrifluoroethane	530	E	0.25	0.50	1.00	ug/L
P5261-09	RE134D3-20241210	Water	Methylene Chloride	0.56	J	0.32	0.50	1.00	ug/L
P5261-09	RE134D3-20241210	Water	cis-1,2-Dichloroethene	2.60		0.25	0.75	1.00	ug/L
P5261-09	RE134D3-20241210	Water	Chloroform	0.31	J	0.26	0.50	1.00	ug/L
P5261-09	RE134D3-20241210	Water	Trichloroethene	130		0.32	0.75	1.00	ug/L
P5261-09	RE134D3-20241210	Water	1,2-Dichloropropane	1.30		0.19	0.50	1.00	ug/L
P5261-09	RE134D3-20241210	Water	Tetrachloroethene	180	E	0.25	0.50	1.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				845					
Total Concentration:				845					
Client ID:	RE134D3-20241210DL								
P5261-09DL	RE134D3-20241210	Water	1,1,2-Trichlorotrifluoroethane	510	D	5.00	10.0	20.0	ug/L
P5261-09DL	RE134D3-20241210	Water	Trichloroethene	150	D	6.40	15.0	20.0	ug/L
P5261-09DL	RE134D3-20241210	Water	Tetrachloroethene	180	D	5.00	10.0	20.0	ug/L
Total Voc :				840					
Total Concentration:				840					
Client ID:	RE134D4-20241210								
P5261-10	RE134D4-20241210	Water	1,1,2-Trichlorotrifluoroethane	450	E	0.25	0.50	1.00	ug/L
P5261-10	RE134D4-20241210	Water	cis-1,2-Dichloroethene	1.20		0.25	0.75	1.00	ug/L
P5261-10	RE134D4-20241210	Water	Trichloroethene	42.4		0.32	0.75	1.00	ug/L
P5261-10	RE134D4-20241210	Water	1,2-Dichloropropane	1.20		0.19	0.50	1.00	ug/L
P5261-10	RE134D4-20241210	Water	Tetrachloroethene	34.9		0.25	0.50	1.00	ug/L
Total Voc :				530					
Total Concentration:				530					
Client ID:	RE134D4-20241210DL								
P5261-10DL	RE134D4-20241210	Water	1,1,2-Trichlorotrifluoroethane	330	D	5.00	10.0	20.0	ug/L
P5261-10DL	RE134D4-20241210	Water	Trichloroethene	43.4	D	6.40	15.0	20.0	ug/L
P5261-10DL	RE134D4-20241210	Water	Tetrachloroethene	28.0	D	5.00	10.0	20.0	ug/L
Total Voc :				401					
Total Concentration:				401					
Client ID:	GM3-RW3-MW1-20241210								
P5261-11	GM3-RW3-MW1-2	Water	Chloroform	0.43	J	0.26	0.50	1.00	ug/L
P5261-11	GM3-RW3-MW1-2	Water	Trichloroethene	18.4		0.32	0.75	1.00	ug/L
P5261-11	GM3-RW3-MW1-2	Water	Tetrachloroethene	1.30		0.25	0.50	1.00	ug/L
Total Voc :				20.1					
Total Concentration:				20.1					
Client ID:	RE125D1-20241210								
P5261-12	RE125D1-20241210	Water	1,1,2-Trichlorotrifluoroethane	2.70		0.25	0.50	1.00	ug/L
P5261-12	RE125D1-20241210	Water	1,1-Dichloroethene	0.77	J	0.26	0.75	1.00	ug/L
P5261-12	RE125D1-20241210	Water	Methyl tert-butyl Ether	0.81	J	0.16	0.50	1.00	ug/L
P5261-12	RE125D1-20241210	Water	1,1-Dichloroethane	0.55	J	0.23	0.50	1.00	ug/L
P5261-12	RE125D1-20241210	Water	cis-1,2-Dichloroethene	2.40		0.25	0.75	1.00	ug/L
P5261-12	RE125D1-20241210	Water	Chloroform	0.78	J	0.26	0.50	1.00	ug/L
P5261-12	RE125D1-20241210	Water	Trichloroethene	200	EM	0.32	0.75	1.00	ug/L
P5261-12	RE125D1-20241210	Water	Tetrachloroethene	11.2		0.25	0.50	1.00	ug/L
Total Voc :				219					
Total Concentration:				219					

Hit Summary Sheet

SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	RE125D1-20241210DL								
P5261-12DL	RE125D1-20241210	Water	1,1,2-Trichlorotrifluoroethane	3.40	JD	1.30	2.50	5.00	ug/L
P5261-12DL	RE125D1-20241210	Water	1,1-Dichloroethene	1.30	JD	1.30	3.80	5.00	ug/L
P5261-12DL	RE125D1-20241210	Water	cis-1,2-Dichloroethene	2.50	JD	1.30	3.80	5.00	ug/L
P5261-12DL	RE125D1-20241210	Water	Trichloroethene	170	D	1.60	3.80	5.00	ug/L
P5261-12DL	RE125D1-20241210	Water	Tetrachloroethene	11.7	D	1.30	2.50	5.00	ug/L
			Total Voc :	189					
			Total Concentration:	189					
Client ID:	RE125D2-20241210								
P5261-15	RE125D2-20241210	Water	1,1,2-Trichlorotrifluoroethane	7.60		0.25	0.50	1.00	ug/L
P5261-15	RE125D2-20241210	Water	1,1-Dichloroethene	3.00		0.26	0.75	1.00	ug/L
P5261-15	RE125D2-20241210	Water	cis-1,2-Dichloroethene	3.90		0.25	0.75	1.00	ug/L
P5261-15	RE125D2-20241210	Water	Chloroform	0.45	J	0.26	0.50	1.00	ug/L
P5261-15	RE125D2-20241210	Water	Trichloroethene	370	E	0.32	0.75	1.00	ug/L
P5261-15	RE125D2-20241210	Water	Tetrachloroethene	13.6		0.25	0.50	1.00	ug/L
			Total Voc :	399					
			Total Concentration:	399					
Client ID:	RE125D2-20241210DL								
P5261-15DL	RE125D2-20241210	Water	1,1,2-Trichlorotrifluoroethane	6.80	JD	2.50	5.00	10.0	ug/L
P5261-15DL	RE125D2-20241210	Water	1,1-Dichloroethene	2.80	JD	2.60	7.50	10.0	ug/L
P5261-15DL	RE125D2-20241210	Water	cis-1,2-Dichloroethene	3.80	JD	2.50	7.50	10.0	ug/L
P5261-15DL	RE125D2-20241210	Water	Trichloroethene	350	D	3.20	7.50	10.0	ug/L
P5261-15DL	RE125D2-20241210	Water	Tetrachloroethene	12.4	D	2.50	5.00	10.0	ug/L
			Total Voc :	376					
			Total Concentration:	376					
Client ID:	RE125D3-20241210								
P5261-16	RE125D3-20241210	Water	1,1,2-Trichlorotrifluoroethane	150	E	0.25	0.50	1.00	ug/L
P5261-16	RE125D3-20241210	Water	cis-1,2-Dichloroethene	1.50		0.25	0.75	1.00	ug/L
P5261-16	RE125D3-20241210	Water	Trichloroethene	79.1		0.32	0.75	1.00	ug/L
P5261-16	RE125D3-20241210	Water	Tetrachloroethene	11.6		0.25	0.50	1.00	ug/L
			Total Voc :	242					
			Total Concentration:	242					
Client ID:	RE125D3-20241210DL								
P5261-16DL	RE125D3-20241210	Water	1,1,2-Trichlorotrifluoroethane	150	D	1.30	2.50	5.00	ug/L
P5261-16DL	RE125D3-20241210	Water	Trichloroethene	81.3	D	1.60	3.80	5.00	ug/L
P5261-16DL	RE125D3-20241210	Water	Tetrachloroethene	10.8	D	1.30	2.50	5.00	ug/L
			Total Voc :	242					
			Total Concentration:	242					
Client ID:	RE120D1-20241210								

Hit Summary Sheet SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
P5261-17	RE120D1-20241210	Water	1,1,2-Trichlorotrifluoroethane	0.88	J	0.25	0.50	1.00	ug/L
P5261-17	RE120D1-20241210	Water	Trichloroethene	11.7		0.32	0.75	1.00	ug/L
			Total Voc :	12.6					
			Total Concentration:	12.6					
Client ID:	RE120D2-20241210								
P5261-18	RE120D2-20241210	Water	1,1,2-Trichlorotrifluoroethane	110		0.25	0.50	1.00	ug/L
P5261-18	RE120D2-20241210	Water	1,1-Dichloroethene	1.80		0.26	0.75	1.00	ug/L
P5261-18	RE120D2-20241210	Water	cis-1,2-Dichloroethene	2.20		0.25	0.75	1.00	ug/L
P5261-18	RE120D2-20241210	Water	Chloroform	0.46	J	0.26	0.50	1.00	ug/L
P5261-18	RE120D2-20241210	Water	Trichloroethene	390	E	0.32	0.75	1.00	ug/L
P5261-18	RE120D2-20241210	Water	Tetrachloroethene	10.1		0.25	0.50	1.00	ug/L
			Total Voc :	515					
			Total Concentration:	515					
Client ID:	RE120D2-20241210DL								
P5261-18DL	RE120D2-20241210	Water	1,1,2-Trichlorotrifluoroethane	100	D	2.50	5.00	10.0	ug/L
P5261-18DL	RE120D2-20241210	Water	cis-1,2-Dichloroethene	2.90	JD	2.50	7.50	10.0	ug/L
P5261-18DL	RE120D2-20241210	Water	Trichloroethene	380	D	3.20	7.50	10.0	ug/L
P5261-18DL	RE120D2-20241210	Water	Tetrachloroethene	9.60	JD	2.50	5.00	10.0	ug/L
			Total Voc :	493					
			Total Concentration:	493					
Client ID:	RE120D3-20241210								
P5261-19	RE120D3-20241210	Water	1,1,2-Trichlorotrifluoroethane	15.1		0.25	0.50	1.00	ug/L
P5261-19	RE120D3-20241210	Water	1,1-Dichloroethene	6.90		0.26	0.75	1.00	ug/L
P5261-19	RE120D3-20241210	Water	cis-1,2-Dichloroethene	3.70		0.25	0.75	1.00	ug/L
P5261-19	RE120D3-20241210	Water	Trichloroethene	800	E	0.32	0.75	1.00	ug/L
P5261-19	RE120D3-20241210	Water	1,1,2-Trichloroethane	0.60	J	0.21	0.50	1.00	ug/L
P5261-19	RE120D3-20241210	Water	Tetrachloroethene	9.10		0.25	0.50	1.00	ug/L
			Total Voc :	835					
			Total Concentration:	835					
Client ID:	RE120D3-20241210DL								
P5261-19DL	RE120D3-20241210	Water	1,1,2-Trichlorotrifluoroethane	14.6	JD	5.00	10.0	20.0	ug/L
P5261-19DL	RE120D3-20241210	Water	1,1-Dichloroethene	8.00	JD	5.20	15.0	20.0	ug/L
P5261-19DL	RE120D3-20241210	Water	Trichloroethene	730	D	6.40	15.0	20.0	ug/L
P5261-19DL	RE120D3-20241210	Water	Tetrachloroethene	12.1	JD	5.00	10.0	20.0	ug/L
			Total Voc :	765					
			Total Concentration:	765					
Client ID:	DUP-02-20241210								
P5261-20	DUP-02-20241210	Water	1,1,2-Trichlorotrifluoroethane	0.85	J	0.25	0.50	1.00	ug/L
P5261-20	DUP-02-20241210	Water	Trichloroethene	11.3		0.32	0.75	1.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				12.2					
Total Concentration:				12.2					