

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5261

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044311.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1216WBS01	Dichlorodifluoromethane	20	19.9	ug/L	100			32	152	
	Chloromethane	20	20.8	ug/L	104			50	139	
	Vinyl chloride	20	19.7	ug/L	99			58	137	
	Bromomethane	20	19.0	ug/L	95			53	141	
	Chloroethane	20	20.3	ug/L	102			60	138	
	Trichlorofluoromethane	20	16.0	ug/L	80			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.1	ug/L	101			70	136	
	1,1-Dichloroethene	20	20.3	ug/L	102			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	19.3	ug/L	97			64	133	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			71	124	
	Methyl Acetate	20	18.9	ug/L	95			56	136	
	Methylene Chloride	20	20.6	ug/L	103			74	124	
	trans-1,2-Dichloroethene	20	20.3	ug/L	102			75	124	
	1,1-Dichloroethane	20	20.4	ug/L	102			77	125	
	Cyclohexane	20	19.4	ug/L	97			71	130	
	2-Butanone	100	98.7	ug/L	99			56	143	
	Carbon Tetrachloride	20	20.4	ug/L	102			72	136	
	cis-1,2-Dichloroethene	20	19.3	ug/L	97			78	123	
	Bromochloromethane	20	16.6	ug/L	83			78	123	
	Chloroform	20	19.6	ug/L	98			79	124	
	1,1,1-Trichloroethane	20	19.8	ug/L	99			74	131	
	Methylcyclohexane	20	20.6	ug/L	103			72	132	
	Benzene	20	20.5	ug/L	103			79	120	
	1,2-Dichloroethane	20	20.0	ug/L	100			73	128	
	Trichloroethene	20	20.4	ug/L	102			79	123	
	1,2-Dichloropropane	20	20.8	ug/L	104			78	122	
	Bromodichloromethane	20	20.8	ug/L	104			79	125	
	4-Methyl-2-Pentanone	100	93.9	ug/L	94			67	130	
	Toluene	20	20.6	ug/L	103			80	121	
	t-1,3-Dichloropropene	20	20.9	ug/L	104			73	127	
	cis-1,3-Dichloropropene	20	21.3	ug/L	106			75	124	
	1,1,2-Trichloroethane	20	20.6	ug/L	103			80	119	
	2-Hexanone	100	98.2	ug/L	98			57	139	
	Dibromochloromethane	20	19.9	ug/L	100			74	126	
	1,2-Dibromoethane	20	20.8	ug/L	104			77	121	
	Tetrachloroethene	20	21.1	ug/L	106			74	129	
	Chlorobenzene	20	20.4	ug/L	102			82	118	
	Ethyl Benzene	20	20.7	ug/L	104			79	121	
	m/p-Xylenes	40	41.6	ug/L	104			80	121	
	o-Xylene	20	20.9	ug/L	104			78	122	
	Styrene	20	20.7	ug/L	104			78	123	
	Bromoform	20	19.1	ug/L	96			66	130	
	Isopropylbenzene	20	21.0	ug/L	105			72	131	
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	21.1	ug/L	106			80	119	
	1,4-Dichlorobenzene	20	20.3	ug/L	102			79	118	
	1,2-Dichlorobenzene	20	20.2	ug/L	101			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5261
Client: AECOM Technical Services, Inc.
Analytical Method: SW8260-Low **Datafile :** VX044311.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1216WBS01	1,2-Dibromo-3-Chloropropane	20	16.8	ug/L	84			62	128	
	1,2,4-Trichlorobenzene	20	20.3	ug/L	102			69	130	
	1,2,3-Trichlorobenzene	20	19.9	ug/L	100			69	129	