

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG NO.: P5261
Lab File ID: VX044308.D Date Analyzed: 12/16/2024
Instrument ID: MSVOA_X Time Analyzed: 08:58
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	139378	5.54	244474	6.75	205865	10.05
UPPER LIMIT	278756	6.038	488948	7.251	411730	10.549
LOWER LIMIT	69689	5.038	122237	6.251	102933	9.549
EPA SAMPLE NO.						
RE122D1-20241209DL	118197	5.55	223567	6.76	195700	10.06
RE122D3-20241209	121193	5.54	224843	6.76	200483	10.05
RE103D1-20241209DL	115522	5.54	214039	6.76	191412	10.06
RE103D2-20241209DL	117251	5.55	214366	6.76	185352	10.06
DUP-01-20241209	107315	5.55	202180	6.76	178520	10.06
RE134D3-20241210DL	115252	5.55	211101	6.76	183791	10.06
RE134D4-20241210DL	120147	5.55	223179	6.76	192838	10.06
GM3-RW3-MW1-20241210	116448	5.55	216541	6.76	187706	10.06
RE125D2-20241210DL	127800	5.55	235549	6.76	201274	10.06
RE125D3-20241210DL	114288	5.54	207751	6.76	184124	10.06
RE120D1-20241210	112654	5.55	207364	6.76	177488	10.06
RE120D2-20241210DL	118530	5.55	215461	6.76	191459	10.06
RE120D3-20241210	111398	5.54	206459	6.76	178249	10.06
DUP-02-20241210	114448	5.55	220754	6.76	195441	10.06
TB	113669	5.54	206160	6.76	175529	10.06
VX1216WBL01	134837	5.54	253656	6.75	222159	10.05
VX1216WBS01	134529	5.54	242066	6.76	212439	10.06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG NO.: P5261
 Lab File ID: VX044308.D Date Analyzed: 12/16/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	139378	5.54	244474	6.75	205865	10.05
UPPER LIMIT	278756	6.038	488948	7.251	411730	10.549
LOWER LIMIT	69689	5.038	122237	6.251	102933	9.549
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.