

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>			Contract:	<u>AECO15</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5261</u>	SAS No.:	<u>P5261</u>	SDG NO.:	<u>P5261</u>
Lab File ID:	<u>VX044308.D</u>			Date Analyzed:	<u>12/16/2024</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>08:58</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS4 AREA #	RT #				
12 HOUR STD	95969	12.018				
UPPER LIMIT	191938	12.518				
LOWER LIMIT	47984.5	11.518				
EPA SAMPLE NO.						
RE122D1-20241209DL	78131	12.02				
RE122D3-20241209	77295	12.02				
RE103D1-20241209DL	78897	12.02				
RE103D2-20241209DL	73736	12.02				
DUP-01-20241209	69896	12.02				
RE134D3-20241210DL	71963	12.02				
RE134D4-20241210DL	77949	12.02				
GM3-RW3-MW1-20241210	76224	12.02				
RE125D2-20241210DL	78713	12.02				
RE125D3-20241210DL	74369	12.02				
RE120D1-20241210	66190	12.02				
RE120D2-20241210DL	73600	12.02				
RE120D3-20241210	70182	12.02				
DUP-02-20241210	77156	12.02				
TB	65733	12.02				
VX1216WBL01	94759	12.02				
VX1216WBS01	96021	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG NO.: P5261
 Lab File ID: VX044308.D Date Analyzed: 12/16/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95969	12.018				
UPPER LIMIT	191938	12.518				
LOWER LIMIT	47984.5	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.