

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15  
 Lab Code: CHEM Case No.: P5261 SAS No.: P5261 SDG NO.: P5261  
 Lab File ID: VX044375.D Date Analyzed: 12/18/2024  
 Instrument ID: MSVOA\_X Time Analyzed: 11:57  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                    | IS4<br>AREA # | RT #   |  |  |  |  |
|--------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD        | 100489        | 12.018 |  |  |  |  |
| UPPER LIMIT        | 200978        | 12.518 |  |  |  |  |
| LOWER LIMIT        | 50244.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO.     |               |        |  |  |  |  |
| RE125D1-20241210   | 70513         | 12.02  |  |  |  |  |
| RE120D3-20241210DL | 83897         | 12.02  |  |  |  |  |
| VX1218WBL01        | 82810         | 12.02  |  |  |  |  |
| VX1218WBS01        | 101950        | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.