

Initial and Continuing Calibration Verification

Client: PARSONS Main of New York, Inc.

SDG No.: P5270

Project: Con Edison Non-MGP - Atlantic Avenue

RunNo.: LB133942

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	11/18/2024
Chloride	mg/L	3	3	100	90-110	11/18/2024
Fluoride	mg/L	2	2	100	90-110	11/18/2024
Nitrite	mg/L	3	3	100	90-110	11/18/2024
Nitrate	mg/L	2.5	2.5	100	90-110	11/18/2024
Sulfate	mg/L	14.6	15	97	90-110	11/18/2024
Orthophosphate as P	mg/L	5	5	100	90-110	11/18/2024
Sample ID: CCV1						
Bromide	mg/L	10.6	10	106	90-110	12/13/2024
Chloride	mg/L	3.1	3	103	90-110	12/13/2024
Fluoride	mg/L	2.1	2	105	90-110	12/13/2024
Nitrite	mg/L	3.2	3	107	90-110	12/13/2024
Nitrate	mg/L	2.6	2.5	104	90-110	12/13/2024
Sulfate	mg/L	15.1	15	101	90-110	12/13/2024
Orthophosphate as P	mg/L	5.2	5	104	90-110	12/13/2024
Sample ID: CCV2						
Bromide	mg/L	10.7	10	107	90-110	12/13/2024
Chloride	mg/L	3.2	3	107	90-110	12/13/2024
Fluoride	mg/L	2.1	2	105	90-110	12/13/2024
Nitrite	mg/L	3.2	3	107	90-110	12/13/2024
Nitrate	mg/L	2.6	2.5	104	90-110	12/13/2024
Sulfate	mg/L	15	15	100	90-110	12/13/2024
Orthophosphate as P	mg/L	5.2	5	104	90-110	12/13/2024