

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270
Instrument ID: BNA_F Calibration Date(s): 11/21/2024 11/21/2024
Calibration Time(s): 11:13 14:18

LAB FILE ID:		RRF2.5 = BF140528.D			RRF005 = BF140529.D		RRF010 = BF140530.D	
		RRF020 = BF140531.D			RRF040 = BF140532.D		RRF050 = BF140533.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.261	1.233	1.202	1.174	1.117	1.172	5.4
Benzaldehyde			1.007	0.970	0.738	0.752	0.820	19.6
Phenol-d6		1.729	1.617	1.559	1.602	1.465	1.550	7.1
Phenol		1.723	1.648	1.620	1.667	1.514	1.583	7.0
bis(2-Chloroethyl) ether		1.292	1.242	1.214	1.246	1.146	1.211	4.6
2-Chlorophenol		1.385	1.328	1.278	1.283	1.201	1.261	6.5
2-Methylphenol		1.121	1.034	1.033	1.042	0.956	1.010	6.7
2,2-oxybis(1-Chloropropane)		1.650	1.480	1.434	1.463	1.335	1.431	8.4
Acetophenone		0.536	0.511	0.494	0.481	0.463	0.488	5.7
3+4-Methylphenols		1.495	1.362	1.305	1.347	1.211	1.298	9.0
n-Nitroso-di-n-propylamine	0.972	1.002	0.949	0.916	0.946	0.856	0.916	6.8
Nitrobenzene-d5		0.409	0.403	0.395	0.392	0.377	0.391	3.2
Hexachloroethane		0.598	0.545	0.549	0.537	0.514	0.536	6.2
Nitrobenzene		0.439	0.409	0.409	0.401	0.388	0.404	4.3
Isophorone		0.694	0.654	0.657	0.662	0.629	0.652	3.4
2-Nitrophenol		0.178	0.172	0.185	0.180	0.178	0.179	2.2
2,4-Dimethylphenol		0.221	0.214	0.213	0.224	0.204	0.214	3.4
bis(2-Chloroethoxy) methane		0.428	0.408	0.403	0.397	0.378	0.397	4.4
2,4-Dichlorophenol		0.301	0.291	0.290	0.282	0.277	0.284	3.8
Naphthalene		1.116	1.075	1.062	1.013	0.993	1.030	5.3
4-Chloroaniline		0.308	0.318	0.309	0.325	0.303	0.308	3.7
Hexachlorobutadiene		0.227	0.225	0.219	0.212	0.209	0.215	4.2
Caprolactam		0.091	0.091	0.091	0.090	0.085	0.088	4.0
4-Chloro-3-methylphenol		0.348	0.318	0.317	0.327	0.307	0.318	5.0
2-Methylnaphthalene		0.724	0.680	0.669	0.650	0.623	0.654	6.1
Hexachlorocyclopentadiene			0.055	0.090	0.114	0.121	0.107	27.8
2,4,6-Trichlorophenol		0.384	0.358	0.375	0.366	0.364	0.367	2.5
2-Fluorobiphenyl		1.550	1.402	1.423	1.291	1.255	1.342	8.9
2,4,5-Trichlorophenol		0.402	0.396	0.410	0.404	0.390	0.399	1.8
1,1-Biphenyl		1.666	1.530	1.563	1.447	1.427	1.493	6.5
2-Chloronaphthalene		1.251	1.145	1.162	1.106	1.082	1.131	5.4
2-Nitroaniline		0.367	0.351	0.379	0.367	0.363	0.363	2.5
Dimethylphthalate		1.455	1.329	1.346	1.299	1.267	1.317	5.3
Acenaphthylene		1.890	1.765	1.808	1.662	1.628	1.710	6.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270
Instrument ID: BNA_F Calibration Date(s): 11/21/2024 11/21/2024
Calibration Time(s): 11:13 14:18

LAB FILE ID:		RRF2.5 = BF140528.D			RRF005 = BF140529.D		RRF010 = BF140530.D	
		RRF020 = BF140531.D			RRF040 = BF140532.D		RRF050 = BF140533.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.314	0.299	0.307	0.301	0.291	0.299	3.2
3-Nitroaniline		0.307	0.297	0.309	0.300	0.289	0.292	5.7
Acenaphthene		1.182	1.110	1.134	1.063	1.052	1.086	5.3
2,4-Dinitrophenol			0.057	0.089	0.140	0.145	0.122	32.7
4-Nitrophenol			0.160	0.195	0.207	0.212	0.199	10.3
Dibenzofuran		1.898	1.739	1.739	1.622	1.559	1.657	8.5
2,4-Dinitrotoluene		0.403	0.403	0.416	0.404	0.386	0.397	3.2
Diethylphthalate		1.495	1.375	1.393	1.311	1.292	1.338	6.7
4-Chlorophenyl-phenylether		0.739	0.682	0.685	0.639	0.619	0.653	7.9
Fluorene		1.509	1.409	1.399	1.295	1.263	1.330	8.4
4-Nitroaniline		0.307	0.301	0.315	0.312	0.310	0.306	2.7
4,6-Dinitro-2-methylphenol			0.073	0.090	0.110	0.110	0.102	16.7
n-Nitrosodiphenylamine		0.632	0.618	0.597	0.578	0.577	0.591	4.4
2,4,6-Tribromophenol		0.222	0.209	0.220	0.216	0.210	0.214	2.5
4-Bromophenyl-phenylether		0.218	0.215	0.206	0.200	0.200	0.206	3.8
Hexachlorobenzene		0.257	0.240	0.239	0.233	0.232	0.238	3.7
Atrazine		0.181	0.171	0.131	0.140	0.147	0.166	16.4
Pentachlorophenol			0.071	0.090	0.116	0.115	0.105	19.2
Phenanthrene		1.074	1.020	0.970	0.944	0.925	0.961	6.9
Anthracene		1.038	0.994	0.958	0.925	0.905	0.940	6.5
Carbazole		1.004	0.949	0.930	0.889	0.876	0.905	6.7
Di-n-butylphthalate		1.144	1.075	1.074	1.023	1.021	1.044	5.6
Fluoranthene		1.186	1.111	1.114	1.014	0.999	1.044	9.1
Pyrene		1.905	1.801	1.897	1.853	1.791	1.849	2.8
Terphenyl-d14		1.351	1.254	1.308	1.283	1.228	1.284	3.3
Butylbenzylphthalate		0.676	0.644	0.694	0.679	0.655	0.666	3.0
3,3-Dichlorobenzidine		0.375	0.383	0.393	0.413	0.406	0.397	4.0
Benzo (a) anthracene		1.409	1.319	1.379	1.286	1.295	1.324	4.3
Chrysene		1.354	1.263	1.218	1.201	1.137	1.211	6.5
Bis (2-ethylhexyl) phthalate		0.904	0.828	0.862	0.833	0.824	0.841	4.0
Di-n-octyl phthalate		1.198	1.133	1.147	1.132	1.147	1.150	2.3
Benzo (b) fluoranthene		1.347	1.256	1.380	1.186	1.248	1.256	6.4
Benzo (k) fluoranthene		1.243	1.236	1.059	1.101	1.022	1.099	9.3
Benzo (a) pyrene		1.062	1.050	1.063	1.007	0.998	1.021	3.8
Indeno (1,2,3-cd) pyrene		1.209	1.283	1.299	1.304	1.306	1.303	4.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270
Instrument ID: BNA_F Calibration Date(s): 11/21/2024 11/21/2024
Calibration Time(s): 11:13 14:18

LAB FILE ID:		RRF2.5 = BF140528.D			RRF005 = BF140529.D		RRF010 = BF140530.D	
		RRF020 = BF140531.D			RRF040 = BF140532.D		RRF050 = BF140533.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.015	1.038	1.063	1.068	1.075	1.067	3.9
Benzo (g,h,i) perylene		1.033	1.090	1.078	1.085	1.094	1.089	3.5
1,2,4,5-Tetrachlorobenzene		0.635	0.604	0.606	0.563	0.561	0.586	5.0
1,4-Dioxane		0.501	0.502	0.576	0.489	0.452	0.493	8.5
2,3,4,6-Tetrachlorophenol		0.307	0.296	0.308	0.312	0.305	0.306	1.7