

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: P5270 SDG NO.: P5270

Lab File ID: BF140871.D

DFTPP Injection Date: 12/17/2024

Instrument ID: BNA_F

DFTPP Injection Time: 08:58

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.9
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.3 (0.8) 1
127	10.0 - 80.0% of mass 198	48.4
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	27.5
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	12.5
442	Greater than 50% of mass 198	80.7
443	15.0 - 24.0% of mass 442	15.3 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140872.D	12/17/2024	09:24
PB165623BL	PB165623BL	BF140875.D	12/17/2024	10:43
PB165623BS	PB165623BS	BF140876.D	12/17/2024	11:09
PB165623BSD	PB165623BSD	BF140877.D	12/17/2024	11:35
MW14-20241212	P5270-08	BF140883.D	12/17/2024	14:19
MW12-20241212DL	P5270-07DL	BF140885.D	12/17/2024	15:12
MW14-20241212DL	P5270-08DL	BF140886.D	12/17/2024	15:38