Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270

Instrument ID: BNA_F Calibration Date(s): 11/21/2024 11/21/2024

Calibration Time(s): 11:13 14:18

LAB FILE ID: RRF2.5 = BF140528.D RRF005 = BF140529.D RRF010 = BF140530.D
RRF020 = BF140531.D RRF040 = BF140532.D RRF050 = BF140533.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.261	1.233	1.202	1.174	1.117	1.172	5.4
Benzaldehyde			1.007	0.970	0.738	0.752	0.820	19.6
Phenol-d6		1.729	1.617	1.559	1.602	1.465	1.550	7.1
Phenol		1.723	1.648	1.620	1.667	1.514	1.583	7.0
bis(2-Chloroethyl)ether		1.292	1.242	1.214	1.246	1.146	1.211	4.6
2-Chlorophenol		1.385	1.328	1.278	1.283	1.201	1.261	6.5
2-Methylphenol		1.121	1.034	1.033	1.042	0.956	1.010	6.7
2,2-oxybis(1-Chloropropane)		1.650	1.480	1.434	1.463	1.335	1.431	8.4
Acetophenone		0.536	0.511	0.494	0.481	0.463	0.488	5.7
3+4-Methylphenols		1.495	1.362	1.305	1.347	1.211	1.298	9.0
n-Nitroso-di-n-propylamine	0.972	1.002	0.949	0.916	0.946	0.856	0.916	6.8
Nitrobenzene-d5		0.409	0.403	0.395	0.392	0.377	0.391	3.2
Hexachloroethane		0.598	0.545	0.549	0.537	0.514	0.536	6.2
Nitrobenzene		0.439	0.409	0.409	0.401	0.388	0.404	4.3
Isophorone		0.694	0.654	0.657	0.662	0.629	0.652	3.4
2-Nitrophenol		0.178	0.172	0.185	0.180	0.178	0.179	2.2
2,4-Dimethylphenol		0.221	0.214	0.213	0.224	0.204	0.214	3.4
bis(2-Chloroethoxy)methane		0.428	0.408	0.403	0.397	0.378	0.397	4.4
2,4-Dichlorophenol		0.301	0.291	0.290	0.282	0.277	0.284	3.8
Naphthalene		1.116	1.075	1.062	1.013	0.993	1.030	5.3
4-Chloroaniline		0.308	0.318	0.309	0.325	0.303	0.308	3.7
Hexachlorobutadiene		0.227	0.225	0.219	0.212	0.209	0.215	4.2
Caprolactam		0.091	0.091	0.091	0.090	0.085	0.088	4.0
4-Chloro-3-methylphenol		0.348	0.318	0.317	0.327	0.307	0.318	5.0
2-Methylnaphthalene		0.724	0.680	0.669	0.650	0.623	0.654	6.1
Hexachlorocyclopentadiene			0.055	0.090	0.114	0.121	0.107	27.8
2,4,6-Trichlorophenol		0.384	0.358	0.375	0.366	0.364	0.367	2.5
2-Fluorobiphenyl		1.550	1.402	1.423	1.291	1.255	1.342	8.9
2,4,5-Trichlorophenol		0.402	0.396	0.410	0.404	0.390	0.399	1.8
1,1-Biphenyl		1.666	1.530	1.563	1.447	1.427	1.493	6.5
2-Chloronaphthalene		1.251	1.145	1.162	1.106	1.082	1.131	5.4
2-Nitroaniline		0.367	0.351	0.379	0.367	0.363	0.363	2.5
Dimethylphthalate		1.455	1.329	1.346	1.299	1.267	1.317	5.3
Acenaphthylene		1.890	1.765	1.808	1.662	1.628	1.710	6.6

All other compounds must meet a minimum RRF of 0.010.

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270

Instrument ID: BNA_F Calibration Date(s): 11/21/2024 11/21/2024

Calibration Time(s): 11:13 14:18

LAB FILE ID: RRF2.5 = BF140528.D RRF005 = BF140529.D RRF010 = BF140530.D
RRF020 = BF140531.D RRF040 = BF140532.D RRF050 = BF140533.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.314	0.299	0.307	0.301	0.291	0.299	3.2
3-Nitroaniline		0.307	0.297	0.309	0.300	0.289	0.292	5.7
Acenaphthene		1.182	1.110	1.134	1.063	1.052	1.086	5.3
2,4-Dinitrophenol			0.057	0.089	0.140	0.145	0.122	32.7
4-Nitrophenol			0.160	0.195	0.207	0.212	0.199	10.3
Dibenzofuran		1.898	1.739	1.739	1.622	1.559	1.657	8.5
2,4-Dinitrotoluene		0.403	0.403	0.416	0.404	0.386	0.397	3.2
Diethylphthalate		1.495	1.375	1.393	1.311	1.292	1.338	6.7
4-Chlorophenyl-phenylether		0.739	0.682	0.685	0.639	0.619	0.653	7.9
Fluorene		1.509	1.409	1.399	1.295	1.263	1.330	8.4
4-Nitroaniline		0.307	0.301	0.315	0.312	0.310	0.306	2.7
4,6-Dinitro-2-methylphenol			0.073	0.090	0.110	0.110	0.102	16.7
n-Nitrosodiphenylamine		0.632	0.618	0.597	0.578	0.577	0.591	4.4
2,4,6-Tribromophenol		0.222	0.209	0.220	0.216	0.210	0.214	2.5
4-Bromophenyl-phenylether		0.218	0.215	0.206	0.200	0.200	0.206	3.8
Hexachlorobenzene		0.257	0.240	0.239	0.233	0.232	0.238	3.7
Atrazine		0.181	0.171	0.131	0.140	0.147	0.166	16.4
Pentachlorophenol			0.071	0.090	0.116	0.115	0.105	19.2
Phenanthrene		1.074	1.020	0.970	0.944	0.925	0.961	6.9
Anthracene		1.038	0.994	0.958	0.925	0.905	0.940	6.5
Carbazole		1.004	0.949	0.930	0.889	0.876	0.905	6.7
Di-n-butylphthalate		1.144	1.075	1.074	1.023	1.021	1.044	5.6
Fluoranthene		1.186	1.111	1.114	1.014	0.999	1.044	9.1
Pyrene		1.905	1.801	1.897	1.853	1.791	1.849	2.8
Terphenyl-d14		1.351	1.254	1.308	1.283	1.228	1.284	3.3
Butylbenzylphthalate		0.676	0.644	0.694	0.679	0.655	0.666	3.0
3,3-Dichlorobenzidine		0.375	0.383	0.393	0.413	0.406	0.397	4.0
Benzo(a)anthracene		1.409	1.319	1.379	1.286	1.295	1.324	4.3
Chrysene		1.354	1.263	1.218	1.201	1.137	1.211	6.5
Bis(2-ethylhexyl)phthalate		0.904	0.828	0.862	0.833	0.824	0.841	4.0
Di-n-octyl phthalate		1.198	1.133	1.147	1.132	1.147	1.150	2.3
Benzo(b)fluoranthene		1.347	1.256	1.380	1.186	1.248	1.256	6.4
Benzo(k)fluoranthene		1.243	1.236	1.059	1.101	1.022	1.099	9.3
Benzo(a)pyrene		1.062	1.050	1.063	1.007	0.998	1.021	3.8
Indeno(1,2,3-cd)pyrene		1.209	1.283	1.299	1.304	1.306	1.303	4.5

All other compounds must meet a minimum RRF of 0.010.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270

Instrument ID: BNA_F Calibration Date(s): 11/21/2024 11/21/2024

Calibration Time(s): 11:13 14:18

LAB FILE ID: RRF2.5 = BF140528.D RRF005 = BF140529.D RRF010 = BF140530.D
RRF020 = BF140531.D RRF040 = BF140532.D RRF050 = BF140533.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		1.015	1.038	1.063	1.068	1.075	1.067	3.9
Benzo(g,h,i)perylene		1.033	1.090	1.078	1.085	1.094	1.089	3.5
1,2,4,5-Tetrachlorobenzene		0.635	0.604	0.606	0.563	0.561	0.586	5.0
1,4-Dioxane		0.501	0.502	0.576	0.489	0.452	0.493	8.5
2,3,4,6-Tetrachlorophenol		0.307	0.296	0.308	0.312	0.305	0.306	1.7

All other compounds must meet a minimum RRF of 0.010.