

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270
Instrument ID: BNA_F Calibration Date(s): 12/16/2024 12/16/2024
Calibration Time(s): 12:13 16:15

LAB FILE ID:		RRF2.5 = BF140855.D			RRF005 = BF140856.D		RRF010 = BF140857.D	
		RRF020 = BF140858.D			RRF040 = BF140859.D		RRF050 = BF140860.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.211	1.236	1.279	1.199	1.231	1.215	4.4
Benzaldehyde		1.048	1.051	0.993	0.864	0.823	0.892	16.3
Phenol-d6		1.692	1.661	1.661	1.567	1.576	1.609	4.5
Phenol		1.756	1.690	1.733	1.630	1.643	1.668	4.4
bis(2-Chloroethyl) ether		1.277	1.242	1.281	1.227	1.237	1.244	2.9
2-Chlorophenol		1.343	1.379	1.357	1.301	1.313	1.325	3.8
2-Methylphenol		1.081	1.105	1.080	1.039	1.039	1.058	3.5
2,2-oxybis(1-Chloropropane)		1.563	1.495	1.519	1.438	1.438	1.465	4.5
Acetophenone		0.526	0.503	0.520	0.493	0.496	0.500	5.4
3+4-Methylphenols		1.503	1.491	1.440	1.355	1.360	1.403	5.8
n-Nitroso-di-n-propylamine	1.005	1.030	1.004	0.999	0.922	0.938	0.965	5.5
Nitrobenzene-d5		0.406	0.401	0.420	0.393	0.397	0.400	4.4
Hexachloroethane		0.600	0.580	0.585	0.556	0.557	0.570	3.9
Nitrobenzene		0.418	0.411	0.425	0.402	0.408	0.410	4.4
Isophorone		0.681	0.672	0.690	0.651	0.664	0.670	3.5
2-Nitrophenol		0.183	0.185	0.193	0.187	0.191	0.189	3.8
2,4-Dimethylphenol		0.221	0.220	0.225	0.220	0.219	0.220	2.6
bis(2-Chloroethoxy) methane		0.431	0.414	0.429	0.417	0.418	0.418	3.0
2,4-Dichlorophenol		0.300	0.310	0.313	0.299	0.301	0.304	3.1
Naphthalene		1.138	1.125	1.126	1.067	1.090	1.099	3.9
4-Chloroaniline		0.371	0.375	0.377	0.361	0.358	0.365	2.7
Hexachlorobutadiene		0.239	0.235	0.244	0.232	0.233	0.235	3.4
Caprolactam		0.085	0.090	0.090	0.090	0.089	0.091	3.9
4-Chloro-3-methylphenol		0.342	0.347	0.347	0.338	0.332	0.342	1.7
2-Methylnaphthalene		0.744	0.734	0.740	0.721	0.700	0.727	2.1
Hexachlorocyclopentadiene			0.091	0.120	0.142	0.163	0.153	30.1
2,4,6-Trichlorophenol		0.348	0.339	0.386	0.364	0.357	0.376	9.7
2-Fluorobiphenyl		1.560	1.410	1.538	1.369	1.345	1.455	5.8
2,4,5-Trichlorophenol		0.413	0.365	0.426	0.397	0.395	0.414	8.3
1,1-Biphenyl		1.678	1.512	1.639	1.513	1.463	1.590	6.5
2-Chloronaphthalene		1.236	1.157	1.254	1.169	1.124	1.217	6.7
2-Nitroaniline		0.350	0.346	0.372	0.358	0.341	0.367	8.1
Dimethylphthalate		1.431	1.367	1.438	1.359	1.308	1.410	5.4
Acenaphthylene		1.874	1.821	1.880	1.734	1.696	1.788	4.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270
Instrument ID: BNA_F Calibration Date(s): 12/16/2024 12/16/2024
Calibration Time(s): 12:13 16:15

LAB FILE ID:		RRF2.5 = BF140855.D			RRF005 = BF140856.D		RRF010 = BF140857.D	
		RRF020 = BF140858.D			RRF040 = BF140859.D		RRF050 = BF140860.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.320	0.317	0.328	0.312	0.303	0.322	5.1
3-Nitroaniline		0.322	0.316	0.329	0.316	0.325	0.320	2.3
Acenaphthene		1.228	1.190	1.166	1.113	1.100	1.142	5.2
2,4-Dinitrophenol			0.088	0.108	0.134	0.145	0.132	21.6
4-Nitrophenol			0.091	0.133	0.165	0.169	0.150	21.6
Dibenzofuran		1.920	1.876	1.830	1.728	1.705	1.770	6.5
2,4-Dinitrotoluene		0.435	0.438	0.440	0.419	0.421	0.424	4.1
Diethylphthalate		1.562	1.484	1.496	1.439	1.446	1.462	4.5
4-Chlorophenyl-phenylether		0.780	0.750	0.759	0.716	0.727	0.733	4.6
Fluorene		1.573	1.522	1.492	1.444	1.429	1.460	5.3
4-Nitroaniline		0.329	0.333	0.339	0.338	0.343	0.334	2.4
4,6-Dinitro-2-methylphenol			0.099	0.109	0.117	0.118	0.113	9.0
n-Nitrosodiphenylamine		0.639	0.679	0.644	0.599	0.592	0.615	7.9
2,4,6-Tribromophenol		0.235	0.236	0.242	0.240	0.235	0.237	2.6
4-Bromophenyl-phenylether		0.225	0.248	0.236	0.220	0.214	0.224	8.0
Hexachlorobenzene		0.257	0.280	0.268	0.246	0.245	0.254	7.2
Atrazine		0.200	0.204	0.193	0.168	0.153	0.172	16.5
Pentachlorophenol			0.062	0.082	0.098	0.102	0.093	18.7
Phenanthrene		1.125	1.081	1.072	0.983	0.988	1.028	6.5
Anthracene		1.091	1.064	1.053	0.970	0.978	1.013	5.7
Carbazole		1.086	1.018	1.026	0.952	0.972	0.992	5.6
Di-n-butylphthalate		1.245	1.206	1.232	1.129	1.185	1.164	8.1
Fluoranthene		1.364	1.291	1.251	1.130	1.138	1.184	11.5
Pyrene		1.904	1.719	1.650	1.699	1.761	1.779	6.1
Terphenyl-d14		1.354	1.247	1.218	1.206	1.212	1.268	5.6
Butylbenzylphthalate		0.650	0.616	0.621	0.646	0.672	0.654	4.6
3,3-Dichlorobenzidine		0.429	0.427	0.449	0.412	0.432	0.428	4.1
Benzo (a) anthracene		1.494	1.473	1.421	1.359	1.388	1.417	3.6
Chrysene		1.361	1.284	1.364	1.271	1.254	1.289	4.4
Bis (2-ethylhexyl) phthalate		0.804	0.807	0.875	0.821	0.855	0.844	4.0
Di-n-octyl phthalate		1.267	1.170	1.348	1.223	1.341	1.277	5.2
Benzo (b) fluoranthene		1.454	1.466	1.450	1.294	1.263	1.388	6.1
Benzo (k) fluoranthene		1.318	1.250	1.223	1.224	1.142	1.191	8.6
Benzo (a) pyrene		1.099	1.084	1.134	1.073	1.056	1.087	3.1
Indeno (1,2,3-cd) pyrene		1.130	1.174	1.330	1.269	1.109	1.248	8.9

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG No.: P5270
Instrument ID: BNA_F Calibration Date(s): 12/16/2024 12/16/2024
Calibration Time(s): 12:13 16:15

LAB FILE ID:		RRF2.5 = BF140855.D			RRF005 = BF140856.D		RRF010 = BF140857.D	
		RRF020 = BF140858.D			RRF040 = BF140859.D		RRF050 = BF140860.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		0.913	0.986	1.092	1.058	0.918	1.034	9.3
Benzo (g,h,i) perylene		0.961	0.972	1.094	1.073	0.944	1.052	9.5
1,2,4,5-Tetrachlorobenzene		0.622	0.573	0.641	0.580	0.576	0.615	6.9
1,4-Dioxane		0.547	0.536	0.537	0.503	0.520	0.523	3.8
2,3,4,6-Tetrachlorophenol		0.297	0.312	0.345	0.354	0.356	0.337	7.0