

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: P5270

SDG NO.: P5270

Lab File ID: BF140837.D

DFTPP Injection Date: 12/13/2024

Instrument ID: BNA_F

DFTPP Injection Time: 11:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.5
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	35.3
70	Less than 2.0% of mass 69	0.3 (0.8) 1
127	10.0 - 80.0% of mass 198	47.4
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	29
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	96.9
443	15.0 - 24.0% of mass 442	18 (18.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140838.D	12/13/2024	12:14
MW12-20241212	P5270-07	BF140849.D	12/13/2024	17:28