

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5270

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN085222.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1217WBS02	Dichlorodifluoromethane	20	27.3	ug/L	137		*	69	116	
	Chloromethane	20	18.8	ug/L	94			65	116	
	Vinyl chloride	20	21.8	ug/L	109			65	117	
	Bromomethane	20	25.2	ug/L	126		*	58	125	
	Chloroethane	20	21.4	ug/L	107			56	128	
	Trichlorofluoromethane	20	21.5	ug/L	108			73	115	
	1,1,2-Trichlorotrifluoroethane	20	20.6	ug/L	103			80	112	
	1,1-Dichloroethene	20	19.4	ug/L	97			74	110	
	Acetone	100	100	ug/L	100			60	125	
	Carbon disulfide	20	19.9	ug/L	100			64	112	
	Methyl tert-butyl Ether	20	18.9	ug/L	95			78	114	
	Methyl Acetate	20	18.7	ug/L	94			67	125	
	Methylene Chloride	20	18.2	ug/L	91			72	114	
	trans-1,2-Dichloroethene	20	19.4	ug/L	97			75	108	
	1,1-Dichloroethane	20	19.2	ug/L	96			78	112	
	Cyclohexane	20	18.9	ug/L	95			75	110	
	2-Butanone	100	94.9	ug/L	95			65	122	
	Carbon Tetrachloride	20	21.9	ug/L	110			77	113	
	cis-1,2-Dichloroethene	20	18.9	ug/L	95			77	110	
	Bromochloromethane	20	18.4	ug/L	92			70	124	
	Chloroform	20	18.9	ug/L	95			79	113	
	1,1,1-Trichloroethane	20	19.4	ug/L	97			80	108	
	Methylcyclohexane	20	20.9	ug/L	104			72	115	
	Benzene	20	20.4	ug/L	102			82	109	
	1,2-Dichloroethane	20	20.2	ug/L	101			80	115	
	Trichloroethene	20	20.6	ug/L	103			77	113	
	1,2-Dichloropropane	20	19.7	ug/L	99			83	111	
	Bromodichloromethane	20	21.0	ug/L	105			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	21.1	ug/L	106			82	110	
	t-1,3-Dichloropropene	20	19.8	ug/L	99			79	110	
	cis-1,3-Dichloropropene	20	20.1	ug/L	101			82	110	
	1,1,2-Trichloroethane	20	20.9	ug/L	104			83	112	
	2-Hexanone	100	110	ug/L	110			73	117	
	Dibromochloromethane	20	21.0	ug/L	105			82	110	
	1,2-Dibromoethane	20	20.8	ug/L	104			81	110	
	Tetrachloroethene	20	20.8	ug/L	104			67	123	
	Chlorobenzene	20	19.1	ug/L	96			82	109	
	Ethyl Benzene	20	19.8	ug/L	99			83	109	
	m/p-Xylenes	40	41.8	ug/L	104			82	110	
	o-Xylene	20	19.9	ug/L	100			83	109	
	Styrene	20	20.3	ug/L	102			80	111	
	Bromoform	20	21.4	ug/L	107			79	109	
	Isopropylbenzene	20	19.7	ug/L	99			83	112	
	1,1,2,2-Tetrachloroethane	20	18.5	ug/L	93			76	118	
	1,3-Dichlorobenzene	20	17.2	ug/L	86			82	108	
	1,4-Dichlorobenzene	20	17.3	ug/L	86			82	107	
	1,2-Dichlorobenzene	20	17.4	ug/L	87			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5270

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN085222.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN1217WBS02	1,2-Dibromo-3-Chloropropane	20	18.6	ug/L	93			68	112	
	1,2,4-Trichlorobenzene	20	15.5	ug/L	78			75	113	
	1,2,3-Trichlorobenzene	20	16.5	ug/L	83			76	114	