

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5270 SAS No.: P5270 SDG NO.: P5270
Lab File ID: VN085189.D Date Analyzed: 12/13/2024
Instrument ID: MSVOA_N Time Analyzed: 11:54
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111343	13.788				
UPPER LIMIT	222686	14.288				
LOWER LIMIT	55671.5	13.288				
EPA SAMPLE NO.						
MW8-20241211	81706	13.79				
MW11-20241211	94301	13.79				
MW15-20241211	86352	13.79				
MW17-20241211	83538	13.79				
MW9-20241212	87781	13.79				
MW13D-20241212	88281	13.79				
MW12-20241212	113617	13.79				
MW16D-20241212	119455	13.79				
TB-20241213	114150	13.79				
VN1213WBL01	89452	13.79				
VN1213WBS01	104811	13.79				
VN1213WBSD01	97109	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.