

Pratik Patel

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Thursday, December 26, 2024 3:42 PM
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: Kumar, Narendra; Brandon-Bazile, Kim; Feranda, Jennifer; Arnone, Russell; Aoanan, Agustin; Leung, Christina (she/her/hers)
Subject: REGION 2 | CASE 51826 | LAB ACE | SDG BH8B0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: P5274-07(BH8B6)C.pdf; P5274-07(BH8B6)B.pdf; P5274-07(BH8B6)A.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon ACE,

Regarding the laboratory's reported issue for Case 51826/SDG BH8B0, please see the Region and QSS input below and let me know if you have any questions.

LAB: "Lab is sending this email with regards to case 51826 and SDG BH8B0.

Lab has analyzed sample BH8B6 for TVOA analysis where previous sample had high concentration TCE and instrument blank was not analyzed after the previous sample. Therefore, Lab has re-analyzed the sample the sample TVOA analysis. However, reanalysis has more than three surrogates outside the QC limits. In this case, Lab has used remaining vials and performed again re-analysis where previous sample had again high concentration of TCE observed. However, all three analysis confirm that there is no any carry over from the previous sample therefore lab would like to confirm that lab will report original analysis for this sample.

Please see attachments for your reference."

REGION: "If consistent with SOW, please advise the laboratory to submit all analyses performed. Thanks."

QSS INPUT: The laboratory's approach is in accordance with the SFAM01.1 SOW guidance described in Exhibit D Section 11.3.8 and 11.4.5. The laboratory is expected to comply with the SOW requirements specified in Exhibit D-Sections-11.3 and 11.4.

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Kumar, Narendra <Kumar.Narendra@epa.gov>

Sent: Thursday, December 26, 2024 2:36 PM

To: Irving, Madison <Madison.Irving@gdit.com>

Cc: Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Aoanan, Agustin <Aoanan.Agustin@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>

Subject: RE: REGION 2 | CASE 51826 | LAB ACE | SDG BH8B0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello Madison,

If consistent with SOW, please advise the laboratory to submit all analyses performed. Thanks.

Narendra Kumar

From: Irving, Madison <Madison.Irving@gdit.com>

Sent: Thursday, December 26, 2024 9:29 AM

To: Kumar, Narendra <Kumar.Narendra@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Aoanan, Agustin <Aoanan.Agustin@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>

Subject: REGION 2 | CASE 51826 | LAB ACE | SDG BH8B0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 2,

Please advise on the laboratory's reported issue below for Case 51826/SDG BH8B0.

LAB: "Lab is sending this email with regards to case 51826 and SDG BH8B0.

Lab has analyzed sample BH8B6 for TVOA analysis where previous sample had high concentration TCE and instrument blank was not analyzed after the previous sample. Therefore, Lab has re-analyzed the sample the sample TVOA analysis. However, reanalysis has more than three surrogates outside the QC limits. In this case, Lab has used remaining vials and performed again re-analysis where previous sample had again high concentration of TCE observed. However, all three analysis confirm that there is no any carry over from the previous sample therefore lab would like to confirm that lab will report original analysis for this sample.

Please see attachments for your reference.”

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport
Sent: Friday, December 20, 2024 9:45 AM
To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 2 | CASE 51826 | LAB ACE | SDGs BH8B0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division
Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Sent: Friday, December 20, 2024 9:21 AM
To: DASsupport <DASsupport@gdit.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 2 | CASE 51826 | LAB ACE | SDGs BH8B0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Lab is sending this email with regards to case 51826 and SDG BH8B0.

Lab has analyzed sample BH8B6 for TVOA analysis where previous sample had high concentration TCE and instrument blank was not analyzed after the previous sample. Therefore, Lab has re-analyzed the sample the sample TVOA analysis. However, reanalysis has more than three surrogates outside the QC limits. In this case, Lab has used remaining vials and performed again re-analysis where previous sample had again high concentration of TCE observed. However, all three analysis confirm that there is no any carry over from the previous sample therefore lab would like to confirm that lab will report original analysis for this sample.

Please see attachments for your reference.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

