

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: P5276

Project: Transfer Station-SPDES

Run No.: LB133962

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133962BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	12/16/2024

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: P5276

Project: Transfer Station-SPDES

Run No.: LB133962

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133962BSD							
Oil and Grease	mg/L	20.0	17.1		86	1	78-114	12/16/2024

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: P5276

Project: Transfer Station-SPDES

Run No.: LB133963

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133963BS							
TSS	mg/L	550	510		93	1	90-110	12/16/2024