

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P5276 SAS No.: P5276 SDG No.: P5276

Instrument ID: MSVOA_N Calibration Date/Time: 12/16/2024 09:42

Lab File ID: VN085207.D Init. Calib. Date(s): 12/06/2024 12/06/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:02 11:37

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.521	1.463	0.5	-3.81	
Toluene	1.750	1.703	0.4	-2.69	
Ethyl Benzene	1.857	1.831	0.1	-1.4	
m/p-Xylenes	0.714	0.703	0.3	-1.54	
o-Xylene	0.671	0.678	0.3	1.04	
1,2-Dichloroethane-d4	2.380	2.398	0.01	0.76	
Toluene-d8	1.407	1.384	0.01	-1.63	
4-Bromofluorobenzene	0.496	0.487	0.2	-1.82	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.