

Surrogate Summary

SDG No.: P5276

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5276-03	001-WILLETS-PT-BLVD(DEC)	1,2-Dichloroethane-d4	30	30.3	101	91	110
		Toluene-d8	30	28.3	94	91	112
		4-Bromofluorobenzene	30	26.6	89	63	112
P5276-04	002-35TH-AVE(DEC)	1,2-Dichloroethane-d4	30	30.9	103	91	110
		Toluene-d8	30	28.3	94	91	112
		4-Bromofluorobenzene	30	26.9	90	63	112
VN1216WBL01	VN1216WBL01	1,2-Dichloroethane-d4	30	31.1	104	91	110
		Toluene-d8	30	29.2	97	91	112
		4-Bromofluorobenzene	30	25.2	84	63	112
VN1216WBS01	VN1216WBS01	1,2-Dichloroethane-d4	30	30.0	100	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	29.7	99	63	112
VN1216WBSD0	VN1216WBSD01	1,2-Dichloroethane-d4	30	30.4	101	91	110
		Toluene-d8	30	29.0	97	91	112
		4-Bromofluorobenzene	30	30.0	100	63	112