

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5276
Client: Tully Environmental, Inc
Analytical Method: SW624.1 **Datafile :** VN085208.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1216WBS01	Benzene	20	18.4	ug/L	92			65	135	
	Toluene	20	18.9	ug/L	95			70	130	
	Ethyl Benzene	20	18.7	ug/L	94			60	140	
	m/p-Xylenes	40	38.0	ug/L	95			87	111	
	o-Xylene	20	19.2	ug/L	96			87	111	