

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5276

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VN085209.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1216WBSD01	Benzene	20	19.5	ug/L	98	6		65	135	20
	Toluene	20	19.5	ug/L	98	3		70	130	20
	Ethyl Benzene	20	19.6	ug/L	98	4		60	140	20
	m/p-Xylenes	40	39.6	ug/L	99	4		87	111	20
	o-Xylene	20	20.3	ug/L	102	6		87	111	20