

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: P5276 SAS No.: P5276 SDG NO.: P5276
Lab File ID: VN085207.D Date Analyzed: 12/16/2024
Instrument ID: MSVOA_N Time Analyzed: 09:42
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	33412	7.81	185962	9.09	170233	11.87
UPPER LIMIT	66824	8.312	371924	9.594	340466	12.365
LOWER LIMIT	16706	7.312	92981	8.594	85116.5	11.365
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (DEC)	27333	7.81	145796	9.10	134362	11.87
002-35TH-AVE (DEC)	23791	7.81	122910	9.10	115669	11.87
VN1216WBL01	28325	7.82	149220	9.10	129681	11.87
VN1216WBS01	30825	7.81	169823	9.10	155663	11.87
VN1216WBSD01	29855	7.81	161867	9.10	148673	11.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.