

## Initial and Continuing Calibration Verification

**Client:** Kleinfelder

**SDG No.:** P5277

**Project:** Tanner G. Duckrey Public School

**RunNo.:** LB133944

| Analyte                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------|-------|--------|------------|------------|------------------------|---------------|
| <b>Sample ID: ICV1</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10     | 10         | 100        | 90-110                 | 11/18/2024    |
| Chloride               | mg/L  | 3      | 3          | 100        | 90-110                 | 11/18/2024    |
| Fluoride               | mg/L  | 2      | 2          | 100        | 90-110                 | 11/18/2024    |
| Nitrite                | mg/L  | 3      | 3          | 100        | 90-110                 | 11/18/2024    |
| Nitrate                | mg/L  | 2.5    | 2.5        | 100        | 90-110                 | 11/18/2024    |
| Sulfate                | mg/L  | 14.6   | 15         | 97         | 90-110                 | 11/18/2024    |
| Orthophosphate as P    | mg/L  | 5      | 5          | 100        | 90-110                 | 11/18/2024    |
| <b>Sample ID: CCV1</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.7   | 10         | 107        | 90-110                 | 12/13/2024    |
| Chloride               | mg/L  | 3.2    | 3          | 107        | 90-110                 | 12/13/2024    |
| Fluoride               | mg/L  | 2.1    | 2          | 105        | 90-110                 | 12/13/2024    |
| Nitrite                | mg/L  | 3.2    | 3          | 107        | 90-110                 | 12/13/2024    |
| Nitrate                | mg/L  | 2.6    | 2.5        | 104        | 90-110                 | 12/13/2024    |
| Sulfate                | mg/L  | 15     | 15         | 100        | 90-110                 | 12/13/2024    |
| Orthophosphate as P    | mg/L  | 5.2    | 5          | 104        | 90-110                 | 12/13/2024    |
| <b>Sample ID: CCV2</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.7   | 10         | 107        | 90-110                 | 12/13/2024    |
| Chloride               | mg/L  | 3.2    | 3          | 107        | 90-110                 | 12/13/2024    |
| Fluoride               | mg/L  | 2.1    | 2          | 105        | 90-110                 | 12/13/2024    |
| Nitrite                | mg/L  | 3.2    | 3          | 107        | 90-110                 | 12/13/2024    |
| Nitrate                | mg/L  | 2.6    | 2.5        | 104        | 90-110                 | 12/13/2024    |
| Sulfate                | mg/L  | 15.2   | 15         | 101        | 90-110                 | 12/13/2024    |
| Orthophosphate as P    | mg/L  | 5.4    | 5          | 108        | 90-110                 | 12/13/2024    |

## Initial and Continuing Calibration Verification

**Client:** Kleinfelder

**SDG No.:** P5277

**Project:** Tanner G. Duckrey Public School

**RunNo.:** LB133973

| Analyte                         | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|---------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: ICV1<br>Ammonia as N | mg/L  | 0.96   | 1          | 96         | 90-110                 | 12/17/2024    |
| Sample ID: CCV1<br>Ammonia as N | mg/L  | 0.98   | 1          | 98         | 90-110                 | 12/17/2024    |
| Sample ID: CCV2<br>Ammonia as N | mg/L  | 0.97   | 1          | 97         | 90-110                 | 12/17/2024    |

## Initial and Continuing Calibration Verification

**Client:** Kleinfelder

**SDG No.:** P5277

**Project:** Tanner G. Duckrey Public School

**RunNo.:** LB133981

| Analyte                                       | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-----------------------------------------------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: <b>ICV</b><br>Hexavalent Chromium  | mg/L  | 0.503  | 0.5        | 101           | 90-110                    | 12/17/2024       |
| Sample ID: <b>CCV1</b><br>Hexavalent Chromium | mg/L  | 0.499  | 0.5        | 100           | 90-110                    | 12/17/2024       |
| Sample ID: <b>CCV2</b><br>Hexavalent Chromium | mg/L  | 0.503  | 0.5        | 101           | 90-110                    | 12/17/2024       |