

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: P5277

Project: Tanner G. Duckrey Public School

RunNo.: LB133944

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	11/18/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/18/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	11/18/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/18/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	11/18/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	11/18/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	11/18/2024
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	12/13/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	12/13/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	12/13/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	12/13/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	12/13/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	12/13/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	12/13/2024
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	12/13/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	12/13/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	12/13/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	12/13/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	12/13/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	12/13/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	12/13/2024

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: P5277

Project: Tanner G. Duckrey Public School

RunNo.: LB133973

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/17/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/17/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/17/2024

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: P5277

Project: Tanner G. Duckrey Public School

RunNo.: LB133981

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	12/17/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	12/17/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	12/17/2024