

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** P5277
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5277 **SAS No.:** P5277
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV19	Mercury	3.88	4.0	97	90 - 110	CV	12/16/2024	11:40	LB133960

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** P5277
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5277 **SAS No.:** P5277
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV29	Mercury	4.83	5.0	97	90 - 110	CV	12/16/2024	11:51	LB133960
CCV30	Mercury	5.32	5.0	106	90 - 110	CV	12/16/2024	14:19	LB133960
CCV31	Mercury	4.69	5.0	94	90 - 110	CV	12/16/2024	15:26	LB133960

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: P5277
Contract: POWE02 Lab Code: CHEM Case No.: P5277 SAS No.: P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2620	2500	105	90 - 110	P	12/16/2024	16:02	LB133968
	Antimony	1020	1000	102	90 - 110	P	12/16/2024	16:02	LB133968
	Arsenic	1030	1000	103	90 - 110	P	12/16/2024	16:02	LB133968
	Barium	520	520	100	90 - 110	P	12/16/2024	16:02	LB133968
	Beryllium	527	510	103	90 - 110	P	12/16/2024	16:02	LB133968
	Boron	2280	2500	91	90 - 110	P	12/16/2024	16:02	LB133968
	Cadmium	519	510	102	90 - 110	P	12/16/2024	16:02	LB133968
	Chromium	540	520	104	90 - 110	P	12/16/2024	16:02	LB133968
	Cobalt	526	520	101	90 - 110	P	12/16/2024	16:02	LB133968
	Copper	541	510	106	90 - 110	P	12/16/2024	16:02	LB133968
	Iron	10200	10000	102	90 - 110	P	12/16/2024	16:02	LB133968
	Lead	1030	1000	103	90 - 110	P	12/16/2024	16:02	LB133968
	Manganese	527	520	101	90 - 110	P	12/16/2024	16:02	LB133968
	Molybdenum	2270	2500	91	90 - 110	P	12/16/2024	16:02	LB133968
	Nickel	529	530	100	90 - 110	P	12/16/2024	16:02	LB133968
	Selenium	1060	1000	106	90 - 110	P	12/16/2024	16:02	LB133968
	Silver	255	250	102	90 - 110	P	12/16/2024	16:02	LB133968
	Thallium	1060	1000	106	90 - 110	P	12/16/2024	16:02	LB133968
	Vanadium	511	500	102	90 - 110	P	12/16/2024	16:02	LB133968
	Zinc	1060	1000	106	90 - 110	P	12/16/2024	16:02	LB133968

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: P5277
Contract: POWE02 Lab Code: CHEM Case No.: P5277 SAS No.: P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	103	100	103	80 - 120	P	12/16/2024	16:06	LB133968
	Antimony	46.4	50.0	93	80 - 120	P	12/16/2024	16:06	LB133968
	Arsenic	19.2	20.0	96	80 - 120	P	12/16/2024	16:06	LB133968
	Barium	95.0	100	95	80 - 120	P	12/16/2024	16:06	LB133968
	Beryllium	5.99	6.0	100	80 - 120	P	12/16/2024	16:06	LB133968
	Boron	110	100	110	80 - 120	P	12/16/2024	16:06	LB133968
	Cadmium	5.53	6.0	92	80 - 120	P	12/16/2024	16:06	LB133968
	Chromium	9.70	10.0	97	80 - 120	P	12/16/2024	16:06	LB133968
	Cobalt	28.0	30.0	93	80 - 120	P	12/16/2024	16:06	LB133968
	Copper	21.6	20.0	108	80 - 120	P	12/16/2024	16:06	LB133968
	Iron	91.1	100	91	80 - 120	P	12/16/2024	16:06	LB133968
	Lead	10.6	12.0	88	80 - 120	P	12/16/2024	16:06	LB133968
	Manganese	20.2	20.0	101	80 - 120	P	12/16/2024	16:06	LB133968
	Molybdenum	196	200	98	80 - 120	P	12/16/2024	16:06	LB133968
	Nickel	37.5	40.0	94	80 - 120	P	12/16/2024	16:06	LB133968
	Selenium	19.7	20.0	98	80 - 120	P	12/16/2024	16:06	LB133968
	Silver	9.45	10.0	94	80 - 120	P	12/16/2024	16:06	LB133968
	Thallium	35.6	40.0	89	80 - 120	P	12/16/2024	16:06	LB133968
	Vanadium	38.7	40.0	97	80 - 120	P	12/16/2024	16:06	LB133968
	Zinc	39.8	40.0	100	80 - 120	P	12/16/2024	16:06	LB133968

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** P5277
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5277 **SAS No.:** P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9630	10000	96	90 - 110	P	12/16/2024	16:57	LB133968
	Antimony	4800	5000	96	90 - 110	P	12/16/2024	16:57	LB133968
	Arsenic	4810	5000	96	90 - 110	P	12/16/2024	16:57	LB133968
	Barium	9520	10000	95	90 - 110	P	12/16/2024	16:57	LB133968
	Beryllium	252	250	101	90 - 110	P	12/16/2024	16:57	LB133968
	Boron	4980	5000	100	90 - 110	P	12/16/2024	16:57	LB133968
	Cadmium	2460	2500	99	90 - 110	P	12/16/2024	16:57	LB133968
	Chromium	993	1000	99	90 - 110	P	12/16/2024	16:57	LB133968
	Cobalt	2440	2500	98	90 - 110	P	12/16/2024	16:57	LB133968
	Copper	1220	1250	97	90 - 110	P	12/16/2024	16:57	LB133968
	Iron	4780	5000	96	90 - 110	P	12/16/2024	16:57	LB133968
	Lead	4930	5000	99	90 - 110	P	12/16/2024	16:57	LB133968
	Manganese	2420	2500	97	90 - 110	P	12/16/2024	16:57	LB133968
	Molybdenum	4810	5000	96	90 - 110	P	12/16/2024	16:57	LB133968
	Nickel	2450	2500	98	90 - 110	P	12/16/2024	16:57	LB133968
	Selenium	4810	5000	96	90 - 110	P	12/16/2024	16:57	LB133968
	Silver	1220	1250	98	90 - 110	P	12/16/2024	16:57	LB133968
	Thallium	4960	5000	99	90 - 110	P	12/16/2024	16:57	LB133968
	Vanadium	2430	2500	97	90 - 110	P	12/16/2024	16:57	LB133968
	Zinc	2460	2500	98	90 - 110	P	12/16/2024	16:57	LB133968
CCV02	Aluminum	9700	10000	97	90 - 110	P	12/16/2024	17:58	LB133968
	Antimony	4720	5000	94	90 - 110	P	12/16/2024	17:58	LB133968
	Arsenic	4740	5000	95	90 - 110	P	12/16/2024	17:58	LB133968
	Barium	9620	10000	96	90 - 110	P	12/16/2024	17:58	LB133968
	Beryllium	263	250	105	90 - 110	P	12/16/2024	17:58	LB133968
	Boron	5130	5000	103	90 - 110	P	12/16/2024	17:58	LB133968
	Cadmium	2500	2500	100	90 - 110	P	12/16/2024	17:58	LB133968
	Chromium	1020	1000	102	90 - 110	P	12/16/2024	17:58	LB133968
	Cobalt	2470	2500	99	90 - 110	P	12/16/2024	17:58	LB133968
	Copper	1220	1250	97	90 - 110	P	12/16/2024	17:58	LB133968
	Iron	4840	5000	97	90 - 110	P	12/16/2024	17:58	LB133968
	Lead	5000	5000	100	90 - 110	P	12/16/2024	17:58	LB133968
	Manganese	2490	2500	99	90 - 110	P	12/16/2024	17:58	LB133968
	Molybdenum	4820	5000	96	90 - 110	P	12/16/2024	17:58	LB133968

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder SDG No.: P5277
Contract: POWE02 Lab Code: CHEM Case No.: P5277 SAS No.: P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Nickel	2480	2500	99	90 - 110	P	12/16/2024	17:58	LB133968
	Selenium	4710	5000	94	90 - 110	P	12/16/2024	17:58	LB133968
	Silver	1240	1250	99	90 - 110	P	12/16/2024	17:58	LB133968
	Thallium	4930	5000	98	90 - 110	P	12/16/2024	17:58	LB133968
	Vanadium	2480	2500	99	90 - 110	P	12/16/2024	17:58	LB133968
	Zinc	2470	2500	99	90 - 110	P	12/16/2024	17:58	LB133968
CCV03	Aluminum	9590	10000	96	90 - 110	P	12/16/2024	18:48	LB133968
	Antimony	4660	5000	93	90 - 110	P	12/16/2024	18:48	LB133968
	Arsenic	4700	5000	94	90 - 110	P	12/16/2024	18:48	LB133968
	Barium	9330	10000	93	90 - 110	P	12/16/2024	18:48	LB133968
	Beryllium	263	250	105	90 - 110	P	12/16/2024	18:48	LB133968
	Boron	5110	5000	102	90 - 110	P	12/16/2024	18:48	LB133968
	Cadmium	2480	2500	99	90 - 110	P	12/16/2024	18:48	LB133968
	Chromium	1020	1000	102	90 - 110	P	12/16/2024	18:48	LB133968
	Cobalt	2450	2500	98	90 - 110	P	12/16/2024	18:48	LB133968
	Copper	1210	1250	96	90 - 110	P	12/16/2024	18:48	LB133968
	Iron	4810	5000	96	90 - 110	P	12/16/2024	18:48	LB133968
	Lead	4950	5000	99	90 - 110	P	12/16/2024	18:48	LB133968
	Manganese	2470	2500	99	90 - 110	P	12/16/2024	18:48	LB133968
	Molybdenum	4760	5000	95	90 - 110	P	12/16/2024	18:48	LB133968
	Nickel	2460	2500	98	90 - 110	P	12/16/2024	18:48	LB133968
	Selenium	4670	5000	93	90 - 110	P	12/16/2024	18:48	LB133968
	Silver	1230	1250	98	90 - 110	P	12/16/2024	18:48	LB133968
	Thallium	4800	5000	96	90 - 110	P	12/16/2024	18:48	LB133968
	Vanadium	2460	2500	98	90 - 110	P	12/16/2024	18:48	LB133968
	Zinc	2440	2500	98	90 - 110	P	12/16/2024	18:48	LB133968
CCV04	Aluminum	9770	10000	98	90 - 110	P	12/16/2024	19:51	LB133968
	Antimony	4690	5000	94	90 - 110	P	12/16/2024	19:51	LB133968
	Arsenic	4740	5000	95	90 - 110	P	12/16/2024	19:51	LB133968
	Barium	9400	10000	94	90 - 110	P	12/16/2024	19:51	LB133968
	Beryllium	274	250	110	90 - 110	P	12/16/2024	19:51	LB133968
	Boron	5280	5000	106	90 - 110	P	12/16/2024	19:51	LB133968
	Cadmium	2550	2500	102	90 - 110	P	12/16/2024	19:51	LB133968
	Chromium	1040	1000	104	90 - 110	P	12/16/2024	19:51	LB133968

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** P5277
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5277 **SAS No.:** P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Cobalt	2510	2500	100	90 - 110	P	12/16/2024	19:51	LB133968
	Copper	1220	1250	98	90 - 110	P	12/16/2024	19:51	LB133968
	Iron	4880	5000	98	90 - 110	P	12/16/2024	19:51	LB133968
	Lead	5050	5000	101	90 - 110	P	12/16/2024	19:51	LB133968
	Manganese	2540	2500	102	90 - 110	P	12/16/2024	19:51	LB133968
	Molybdenum	4860	5000	97	90 - 110	P	12/16/2024	19:51	LB133968
	Nickel	2510	2500	101	90 - 110	P	12/16/2024	19:51	LB133968
	Selenium	4680	5000	94	90 - 110	P	12/16/2024	19:51	LB133968
	Silver	1240	1250	100	90 - 110	P	12/16/2024	19:51	LB133968
	Thallium	4970	5000	99	90 - 110	P	12/16/2024	19:51	LB133968
	Vanadium	2530	2500	101	90 - 110	P	12/16/2024	19:51	LB133968
	Zinc	2460	2500	98	90 - 110	P	12/16/2024	19:51	LB133968
CCV05	Aluminum	9980	10000	100	90 - 110	P	12/16/2024	20:49	LB133968
	Antimony	4980	5000	100	90 - 110	P	12/16/2024	20:49	LB133968
	Arsenic	4960	5000	99	90 - 110	P	12/16/2024	20:49	LB133968
	Barium	9720	10000	97	90 - 110	P	12/16/2024	20:49	LB133968
	Beryllium	263	250	105	90 - 110	P	12/16/2024	20:49	LB133968
	Boron	5200	5000	104	90 - 110	P	12/16/2024	20:49	LB133968
	Cadmium	2520	2500	101	90 - 110	P	12/16/2024	20:49	LB133968
	Chromium	1030	1000	103	90 - 110	P	12/16/2024	20:49	LB133968
	Cobalt	2500	2500	100	90 - 110	P	12/16/2024	20:49	LB133968
	Copper	1260	1250	101	90 - 110	P	12/16/2024	20:49	LB133968
	Iron	4950	5000	99	90 - 110	P	12/16/2024	20:49	LB133968
	Lead	5060	5000	101	90 - 110	P	12/16/2024	20:49	LB133968
	Manganese	2500	2500	100	90 - 110	P	12/16/2024	20:49	LB133968
	Molybdenum	5000	5000	100	90 - 110	P	12/16/2024	20:49	LB133968
	Nickel	2510	2500	100	90 - 110	P	12/16/2024	20:49	LB133968
	Selenium	4960	5000	99	90 - 110	P	12/16/2024	20:49	LB133968
	Silver	1270	1250	101	90 - 110	P	12/16/2024	20:49	LB133968
	Thallium	5030	5000	101	90 - 110	P	12/16/2024	20:49	LB133968
	Vanadium	2530	2500	101	90 - 110	P	12/16/2024	20:49	LB133968
	Zinc	2520	2500	101	90 - 110	P	12/16/2024	20:49	LB133968
CCV06	Aluminum	9700	10000	97	90 - 110	P	12/16/2024	21:47	LB133968
	Antimony	4710	5000	94	90 - 110	P	12/16/2024	21:47	LB133968

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** P5277
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5277 **SAS No.:** P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Arsenic	4730	5000	95	90 - 110	P	12/16/2024	21:47	LB133968
	Barium	9350	10000	94	90 - 110	P	12/16/2024	21:47	LB133968
	Beryllium	268	250	107	90 - 110	P	12/16/2024	21:47	LB133968
	Boron	5220	5000	104	90 - 110	P	12/16/2024	21:47	LB133968
	Cadmium	2510	2500	100	90 - 110	P	12/16/2024	21:47	LB133968
	Chromium	1030	1000	103	90 - 110	P	12/16/2024	21:47	LB133968
	Cobalt	2470	2500	99	90 - 110	P	12/16/2024	21:47	LB133968
	Copper	1220	1250	97	90 - 110	P	12/16/2024	21:47	LB133968
	Iron	4840	5000	97	90 - 110	P	12/16/2024	21:47	LB133968
	Lead	5020	5000	100	90 - 110	P	12/16/2024	21:47	LB133968
	Manganese	2490	2500	100	90 - 110	P	12/16/2024	21:47	LB133968
	Molybdenum	4830	5000	96	90 - 110	P	12/16/2024	21:47	LB133968
	Nickel	2480	2500	99	90 - 110	P	12/16/2024	21:47	LB133968
	Selenium	4700	5000	94	90 - 110	P	12/16/2024	21:47	LB133968
	Silver	1250	1250	100	90 - 110	P	12/16/2024	21:47	LB133968
	Thallium	4870	5000	97	90 - 110	P	12/16/2024	21:47	LB133968
	Vanadium	2490	2500	100	90 - 110	P	12/16/2024	21:47	LB133968
	Zinc	2470	2500	99	90 - 110	P	12/16/2024	21:47	LB133968
CCV07	Aluminum	9690	10000	97	90 - 110	P	12/16/2024	23:02	LB133968
	Antimony	4630	5000	92	90 - 110	P	12/16/2024	23:02	LB133968
	Arsenic	4700	5000	94	90 - 110	P	12/16/2024	23:02	LB133968
	Barium	9330	10000	93	90 - 110	P	12/16/2024	23:02	LB133968
	Beryllium	232	250	93	90 - 110	P	12/16/2024	23:02	LB133968
	Boron	5340	5000	107	90 - 110	P	12/16/2024	23:02	LB133968
	Cadmium	2560	2500	102	90 - 110	P	12/16/2024	23:02	LB133968
	Chromium	1050	1000	105	90 - 110	P	12/16/2024	23:02	LB133968
	Cobalt	2500	2500	100	90 - 110	P	12/16/2024	23:02	LB133968
	Copper	1210	1250	96	90 - 110	P	12/16/2024	23:02	LB133968
	Iron	4820	5000	96	90 - 110	P	12/16/2024	23:02	LB133968
	Lead	5100	5000	102	90 - 110	P	12/16/2024	23:02	LB133968
	Manganese	2520	2500	101	90 - 110	P	12/16/2024	23:02	LB133968
	Molybdenum	4790	5000	96	90 - 110	P	12/16/2024	23:02	LB133968
	Nickel	2520	2500	101	90 - 110	P	12/16/2024	23:02	LB133968
	Selenium	4640	5000	93	90 - 110	P	12/16/2024	23:02	LB133968

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Kleinfelder **SDG No.:** P5277
Contract: POWE02 **Lab Code:** CHEM **Case No.:** P5277 **SAS No.:** P5277
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1250	1250	100	90 - 110	P	12/16/2024	23:02	LB133968
	Thallium	4980	5000	100	90 - 110	P	12/16/2024	23:02	LB133968
	Vanadium	2510	2500	100	90 - 110	P	12/16/2024	23:02	LB133968
	Zinc	2460	2500	99	90 - 110	P	12/16/2024	23:02	LB133968
CCV08	Aluminum	9720	10000	97	90 - 110	P	12/16/2024	23:35	LB133968
	Antimony	4760	5000	95	90 - 110	P	12/16/2024	23:35	LB133968
	Arsenic	4800	5000	96	90 - 110	P	12/16/2024	23:35	LB133968
	Barium	9410	10000	94	90 - 110	P	12/16/2024	23:35	LB133968
	Beryllium	274	250	109	90 - 110	P	12/16/2024	23:35	LB133968
	Boron	5270	5000	105	90 - 110	P	12/16/2024	23:35	LB133968
	Cadmium	2540	2500	102	90 - 110	P	12/16/2024	23:35	LB133968
	Chromium	1050	1000	105	90 - 110	P	12/16/2024	23:35	LB133968
	Cobalt	2500	2500	100	90 - 110	P	12/16/2024	23:35	LB133968
	Copper	1230	1250	99	90 - 110	P	12/16/2024	23:35	LB133968
	Iron	4830	5000	97	90 - 110	P	12/16/2024	23:35	LB133968
	Lead	5070	5000	102	90 - 110	P	12/16/2024	23:35	LB133968
	Manganese	2500	2500	100	90 - 110	P	12/16/2024	23:35	LB133968
	Molybdenum	4850	5000	97	90 - 110	P	12/16/2024	23:35	LB133968
	Nickel	2520	2500	101	90 - 110	P	12/16/2024	23:35	LB133968
	Selenium	4760	5000	95	90 - 110	P	12/16/2024	23:35	LB133968
	Silver	1260	1250	101	90 - 110	P	12/16/2024	23:35	LB133968
	Thallium	4900	5000	98	90 - 110	P	12/16/2024	23:35	LB133968
	Vanadium	2510	2500	100	90 - 110	P	12/16/2024	23:35	LB133968
	Zinc	2490	2500	100	90 - 110	P	12/16/2024	23:35	LB133968