

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** P5277 **SAS No.:** P5277 **SDG NO.:** P5277

Continuing Calib Date: 12/16/2024 **Initial Calibration Date(s):** 12/06/2024 12/06/2024

Continuing Calib Time: 16:45 **Initial Calibration Time(s):** 14:19 22:34

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl	8.78	8.79	8.69	8.89	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** P5277 **SAS No.:** P5277 **SDG NO.:** P5277

Continuing Calib Date: 12/16/2024 **Initial Calibration Date(s):** 12/06/2024 12/06/2024

Continuing Calib Time: 16:45 **Initial Calibration Time(s):** 14:19 22:34

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.01
Decachlorobiphenyl	8.73	8.74	8.64	8.84	0.01