



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5277

Client: Kleinfelder

Analytical Method: **8082A** **Datafile :** PO108546.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165646BS	AR1016	166.5	162	ug/kg	97				71	120	
	AR1260	166.5	168	ug/kg	101				65	130	