

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 11/25/2024 11/25/2024

Continuing Calib Time: 10:35 Initial Calibration Time(s): 11:32 12:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.55	3.54	3.44	3.64	-0.01
Aldrin	5.27	5.26	5.16	5.36	-0.01
Dieldrin	6.35	6.34	6.24	6.44	-0.01
4,4'-DDE	6.20	6.19	6.09	6.29	-0.01
4,4'-DDD	6.72	6.71	6.61	6.81	-0.01
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** P5277 **SAS No.:** P5277 **SDG NO.:** P5277

Continuing Calib Date: 12/16/2024 **Initial Calibration Date(s):** 11/25/2024 11/25/2024

Continuing Calib Time: 10:35 **Initial Calibration Time(s):** 11:32 12:25

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.78	2.77	2.67	2.87	-0.01
Aldrin	4.23	4.23	4.13	4.33	0.00
Dieldrin	5.36	5.36	5.26	5.46	0.00
4,4'-DDE	5.23	5.23	5.13	5.33	0.00
4,4'-DDD	5.79	5.79	5.69	5.89	0.00
4,4'-DDT	6.04	6.04	5.94	6.14	0.00