

### CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL01 Date Analyzed: 12/16/2024

Lab Sample No.: PSTDCCC050 Data File : PL093381.D Time Analyzed: 10:35

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| 4,4'-DDD             | 6.717 | 6.608     | 6.808 | 51.220             | 50.000            | 2.4  |
| 4,4'-DDE             | 6.199 | 6.090     | 6.290 | 50.020             | 50.000            | 0.0  |
| 4,4'-DDT             | 7.031 | 6.922     | 7.122 | 49.370             | 50.000            | -1.3 |
| Aldrin               | 5.266 | 5.155     | 5.355 | 50.780             | 50.000            | 1.6  |
| Decachlorobiphenyl   | 9.061 | 8.954     | 9.154 | 55.720             | 50.000            | 11.4 |
| Dieldrin             | 6.352 | 6.242     | 6.442 | 50.460             | 50.000            | 0.9  |
| Tetrachloro-m-xylene | 3.547 | 3.436     | 3.636 | 50.930             | 50.000            | 1.9  |

### CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL01 Date Analyzed: 12/16/2024

Lab Sample No.: PSTDCCC050 Data File : PL093381.D Time Analyzed: 10:35

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| 4,4'-DDD             | 5.788 | 5.685     | 5.885 | 53.710             | 50.000            | 7.4  |
| 4,4'-DDE             | 5.233 | 5.130     | 5.330 | 53.120             | 50.000            | 6.2  |
| 4,4'-DDT             | 6.038 | 5.935     | 6.135 | 52.470             | 50.000            | 4.9  |
| Aldrin               | 4.228 | 4.125     | 4.325 | 54.270             | 50.000            | 8.5  |
| Decachlorobiphenyl   | 7.914 | 7.812     | 8.012 | 52.810             | 50.000            | 5.6  |
| Dieldrin             | 5.364 | 5.262     | 5.462 | 55.020             | 50.000            | 10.0 |
| Tetrachloro-m-xylene | 2.776 | 2.674     | 2.874 | 52.640             | 50.000            | 5.3  |