

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL02 Date Analyzed: 12/16/2024

Lab Sample No.: PSTDCCC050 Data File : PL093393.D Time Analyzed: 16:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.715	6.608	6.808	54.550	50.000	9.1
4,4'-DDE	6.197	6.090	6.290	50.770	50.000	1.5
4,4'-DDT	7.028	6.922	7.122	42.130	50.000	-15.7
Aldrin	5.263	5.155	5.355	51.480	50.000	3.0
Decachlorobiphenyl	9.059	8.954	9.154	55.780	50.000	11.6
Dieldrin	6.349	6.242	6.442	51.060	50.000	2.1
Tetrachloro-m-xylene	3.545	3.436	3.636	51.960	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/25/2024 11/25/2024

Client Sample No.: CCAL02 Date Analyzed: 12/16/2024

Lab Sample No.: PSTDCCC050 Data File : PL093393.D Time Analyzed: 16:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.787	5.685	5.885	59.220	50.000	18.4
4,4'-DDE	5.232	5.130	5.330	54.650	50.000	9.3
4,4'-DDT	6.037	5.935	6.135	43.300	50.000	-13.4
Aldrin	4.227	4.125	4.325	55.890	50.000	11.8
Decachlorobiphenyl	7.913	7.812	8.012	55.030	50.000	10.1
Dieldrin	5.364	5.262	5.462	55.860	50.000	11.7
Tetrachloro-m-xylene	2.776	2.674	2.874	54.540	50.000	9.1