

Final Vol:

11/25/24

10000

uL



Report of Analysis

Client: Kleinfelder Date Collected:

Units:

Project: Tanner G. Duckrey Public School Date Received: 11/25/24

Client Sample ID: PIBLK-PL093230.D SDG No.: P5277

Lab Sample ID: I.BLK-PL093230.D Matrix: WATER

Analytical Method: SW8081 % Solid: 0 Decanted:

Soil Aliquot Vol: uL Test: PESTICIDE Group1

Extraction Type: Injection Volume:

mL

GPC Factor: 1.0 PH:

1000

Prep Method: 3510C

Sample Wt/Vol:

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID
PL093230.D 1 11/25/24 PL112524

Qualifier MDL LOQ / CRQL Units **CAS Number** Parameter Conc. **TARGETS** 309-00-2 Aldrin 0.0044 U 0.0044 0.050 ug/L 60-57-1 Dieldrin 0.0047 U 0.0047 0.050 ug/L 4,4-DDE 0.0045 U 72-55-9 0.0045 0.050 ug/L 72-54-8 4,4-DDD 0.0092 U 0.0092 0.050 ug/L 50-29-3 4,4-DDT 0.0044U 0.00440.050 ug/L **SURROGATES** 2051-24-3 Decachlorobiphenvl 21.6 43 - 140 108% SPK: 20 877-09-8 Tetrachloro-m-xylene 21.2 77 - 126 106% SPK: 20

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit