

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121624\
 Data File : BF140859.D
 Acq On : 16 Dec 2024 14:26
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Quant Time: Dec 16 16:49:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 16 16:38:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.851	152	77674	20.000	ng	0.00
21) Naphthalene-d8	8.128	136	293537	20.000	ng	0.00
39) Acenaphthene-d10	9.886	164	179301	20.000	ng	0.00
64) Phenanthrene-d10	11.375	188	361848	20.000	ng	0.00
76) Chrysene-d12	14.033	240	241027	20.000	ng	0.00
86) Perylene-d12	15.545	264	202727	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	372670	78.982	ng	0.00
7) Phenol-d6	6.498	99	486855	77.902	ng	0.00
23) Nitrobenzene-d5	7.416	82	461011	78.576	ng	0.00
42) 2,4,6-Tribromophenol	10.681	330	172188	81.197	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	981889	75.290	ng	0.00
79) Terphenyl-d14	12.957	244	1163098	76.093	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.622	88	78111	38.433	ng	100
3) Pyridine	3.405	79	189803	40.326	ng	100
4) n-Nitrosodimethylamine	3.357	42	93747	39.510	ng	100
6) Aniline	6.516	93	213440	39.944	ng	100
8) 2-Chlorophenol	6.646	128	202098	39.277	ng	100
9) Benzaldehyde	6.410	77	134222	38.761	ng	100
10) Phenol	6.510	94	253259	39.100	ng	100
11) bis(2-Chloroethyl)ether	6.587	93	190608	39.459	ng	100
12) 1,3-Dichlorobenzene	6.787	146	228181	38.918	ng	100
13) 1,4-Dichlorobenzene	6.869	146	232493	39.065	ng	100
14) 1,2-Dichlorobenzene	7.022	146	223134	39.737	ng	100
15) Benzyl Alcohol	6.998	79	178597	39.968	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.122	45	223455	39.269	ng	100
17) 2-Methylphenol	7.116	107	161380	39.284	ng	100
18) Hexachloroethane	7.357	117	86395	38.994	ng	100
19) n-Nitroso-di-n-propyla...	7.263	70	143279	38.220	ng	100
20) 3+4-Methylphenols	7.269	107	210531	38.648	ng	100
22) Acetophenone	7.263	105	289650	39.446	ng	100
24) Nitrobenzene	7.434	77	235847	39.238	ng	100
25) Isophorone	7.675	82	381938	38.852	ng	100
26) 2-Nitrophenol	7.751	139	109806	39.666	ng	100
27) 2,4-Dimethylphenol	7.793	122	128970	39.892	ng	100
28) bis(2-Chloroethoxy)met...	7.875	93	244652	39.846	ng	100
29) 2,4-Dichlorophenol	7.998	162	175409	39.299	ng	100
30) 1,2,4-Trichlorobenzene	8.069	180	200471	39.180	ng	100
31) Naphthalene	8.151	128	626585	38.860	ng	100
32) Benzoic acid	7.940	122	124210	40.510	ng	100
33) 4-Chloroaniline	8.204	127	211976	39.582	ng	100
34) Hexachlorobutadiene	8.263	225	136482	39.646	ng	100
35) Caprolactam	8.587	113	53059	39.930	ng	100
36) 4-Chloro-3-methylphenol	8.698	107	198298	39.473	ng	100
37) 2-Methylnaphthalene	8.839	142	423369	39.693	ng	100
38) 1-Methylnaphthalene	8.939	142	414363	39.281	ng	100
40) 1,2,4,5-Tetrachloroben...	9.010	216	208141	37.768	ng	100
41) Hexachlorocyclopentadiene	8.987	237	50793	39.476	ng	100
43) 2,4,6-Trichlorophenol	9.128	196	130363	38.631	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121624\
 Data File : BF140859.D
 Acq On : 16 Dec 2024 14:26
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Quant Time: Dec 16 16:49:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 16 16:38:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.181	196	142411	38.395	ng	100
46) 1,1'-Biphenyl	9.304	154	542419	38.044	ng	100
47) 2-Chloronaphthalene	9.334	162	419357	38.447	ng	100
48) 2-Nitroaniline	9.434	65	128421	38.984	ng	100
49) Acenaphthylene	9.745	152	621763	38.790	ng	100
50) Dimethylphthalate	9.604	163	487517	38.566	ng	100
51) 2,6-Dinitrotoluene	9.675	165	111990	38.752	ng	100
52) Acenaphthene	9.916	154	398965	38.965	ng	100
53) 3-Nitroaniline	9.851	138	113227	39.481	ng	100
54) 2,4-Dinitrophenol	9.969	184	47982	37.788	ng	100
55) Dibenzofuran	10.092	168	619833	39.066	ng	100
56) 4-Nitrophenol	10.034	139	59043	40.590	ng	100
57) 2,4-Dinitrotoluene	10.086	165	150199	39.502	ng	100
58) Fluorene	10.434	166	517920	39.568	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.216	232	127023	42.028	ng	100
60) Diethylphthalate	10.304	149	516146	39.375	ng	100
61) 4-Chlorophenyl-phenyle...	10.422	204	256836	39.062	ng	100
62) 4-Nitroaniline	10.469	138	121143	40.423	ng	100
63) Azobenzene	10.581	77	490731	39.821	ng	100
65) 4,6-Dinitro-2-methylph...	10.498	198	84370	41.148	ng	100
66) n-Nitrosodiphenylamine	10.545	169	433296	38.947	ng	100
67) 4-Bromophenyl-phenylether	10.910	248	159117	39.307	ng	100
68) Hexachlorobenzene	10.986	284	177811	38.742	ng	100
69) Atrazine	11.069	200	121493	39.101	ng	100
70) Pentachlorophenol	11.186	266	70638	42.072	ng	100
71) Phenanthrene	11.398	178	711714	38.248	ng	100
72) Anthracene	11.451	178	702321	38.326	ng	100
73) Carbazole	11.610	167	689096	38.392	ng	100
74) Di-n-butylphthalate	11.928	149	816942	38.782	ng	100
75) Fluoranthene	12.592	202	817749	38.159	ng	100
77) Benzidine	12.716	184	194791	39.598	ng	100
78) Pyrene	12.822	202	818937	38.208	ng	100
80) Butylbenzylphthalate	13.427	149	311465	39.493	ng	100
81) Benzo(a)anthracene	14.022	228	654937	38.340	ng	100
82) 3,3'-Dichlorobenzidine	13.980	252	198481	38.511	ng	100
83) Chrysene	14.063	228	612484	39.430	ng	100
84) Bis(2-ethylhexyl)phtha...	13.992	149	395970	38.939	ng	100
85) Di-n-octyl phthalate	14.633	149	589634	38.326	ng	100
87) Indeno(1,2,3-cd)pyrene	17.051	276	514416	40.662	ng	100
88) Benzo(b)fluoranthene	15.110	252	524660	37.283	ng	100
89) Benzo(k)fluoranthene	15.139	252	496310	41.115	ng	100
90) Benzo(a)pyrene	15.480	252	435015	39.478	ng	100
91) Dibenzo(a,h)anthracene	17.057	278	428983	40.931	ng	100
92) Benzo(g,h,i)perylene	17.504	276	434898	40.771	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121624\
Data File : BF140859.D
Acq On : 16 Dec 2024 14:26
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Dec 16 16:49:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 16 16:38:27 2024
Response via : Initial Calibration

