

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5277

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P5277-01MSD	Client Sample ID:	COMP-1AMSD					DataFile:	BF140910.D		
Naphthalene	1900	0	1700	ug/Kg	89	7			72	110	20
Fluorene	1900	0	1800	ug/Kg	95	0			68	116	20
Phenanthrene	1900	0	1900	ug/Kg	100	0			52	128	20
Anthracene	1900	0	2000	ug/Kg	105	0			62	124	20
Pyrene	1900	0	2000	ug/Kg	105	6			26	142	20
Benzo(a)anthracene	1900	0	1900	ug/Kg	100	5			71	114	20
Chrysene	1900	0	1700	ug/Kg	89	12			57	121	20
Benzo(b)fluoranthene	1900	0	2000	ug/Kg	105	10			67	121	20
Benzo(a)pyrene	1900	0	1900	ug/Kg	100	5			70	142	20
Indeno(1,2,3-cd)pyrene	1900	0	2100	ug/Kg	111	6			40	129	20
Benzo(g,h,i)perylene	1900	0	1600	ug/Kg	84	12			24	125	20