

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: P5277

SDG NO.: P5277

Lab File ID: BF140895.D

DFTPP Injection Date: 12/18/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.3
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	34.8
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	47.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	29.2
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	14.2
442	Greater than 50% of mass 198	95.6
443	15.0 - 24.0% of mass 442	17.6 (18.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140896.D	12/18/2024	10:43
PB165648BL	PB165648BL	BF140897.D	12/18/2024	11:27
PB165648BS	PB165648BS	BF140898.D	12/18/2024	11:53
COMP-2A	P5277-02	BF140905.D	12/18/2024	15:02
COMP-3A	P5277-03	BF140906.D	12/18/2024	15:28
COMP-1A	P5277-01	BF140908.D	12/18/2024	16:20
COMP-1AMS	P5277-01MS	BF140909.D	12/18/2024	16:46
COMP-1AMSD	P5277-01MSD	BF140910.D	12/18/2024	17:12