

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121724

Review By	Maresh Dadoda	Review On	12/18/2024 8:38:34 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/18/2024 8:51:04 AM
SubDirectory	VY121724	HP Acquire Method	HP Processing Method 82y121724s.m

STD. NAME	STD REF.#
Tune/Reschk	VP132182
Initial Calibration Stds	VP132183,VP132184,VP132185,VP132186,VP132187,VP132188
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP132189
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020611.D	17 Dec 2024 09:00		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY020612.D	17 Dec 2024 09:38		SY/MD	Ok
3	VSTDICC010	VSTDICC010	VY020613.D	17 Dec 2024 10:01	Com #16,30 on LR	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY020614.D	17 Dec 2024 10:23	Com #11 on QR	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY020615.D	17 Dec 2024 10:51	Com #13 fail for method	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY020616.D	17 Dec 2024 11:14		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY020617.D	17 Dec 2024 11:37		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020618.D	17 Dec 2024 12:24		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY020619.D	17 Dec 2024 14:51	%D fail for com #11	SY/MD	Ok,M
10	VSTDICV050	ICVVY121724	VY020620.D	17 Dec 2024 17:41	ICV fail for com #11,16	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121824

Review By	Maresh Dadoda	Review On	12/19/2024 2:48:41 PM
Supervise By	Semsettin Yesilyurt	Supervise On	12/19/2024 2:49:14 PM
SubDirectory	VY121824	HP Acquire Method	HP Processing Method 82y121724s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132202
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132203,VP132204 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020622.D	18 Dec 2024 08:46		SY/MD	Ok
2	VSTDICV050	ICVVY121824	VY020623.D	18 Dec 2024 09:43		SY/MD	Ok,M
3	VSTDCCC050	VSTDCCC050	VY020624.D	18 Dec 2024 10:06		SY/MD	Ok,M
4	VY1218SBL01	VY1218SBL01	VY020625.D	18 Dec 2024 10:52		SY/MD	Ok
5	VY1218SBS01	VY1218SBS01	VY020626.D	18 Dec 2024 11:19		SY/MD	Ok,M
6	VY1218SBSD01	VY1218SBSD01	VY020627.D	18 Dec 2024 11:42		SY/MD	Ok,M
7	P5277-01	COMP-1A	VY020628.D	18 Dec 2024 12:05	vial-A Not purged	SY/MD	Not Ok
8	P5277-02	COMP-2A	VY020629.D	18 Dec 2024 12:33	vial-A	SY/MD	Ok
9	P5277-03	COMP-3A	VY020630.D	18 Dec 2024 12:57	vial-A	SY/MD	Ok
10	P5287-01	STONES-A	VY020631.D	18 Dec 2024 13:20	vial-A	SY/MD	Ok
11	P5287-03	STONES-B	VY020632.D	18 Dec 2024 13:44	vial-A	SY/MD	Ok
12	P5312-03	CONCRETE-VNJ-222	VY020633.D	18 Dec 2024 14:07	vial-A Surrogate Fail	SY/MD	ReRun
13	P5312-01	SOIL-VNJ-222	VY020634.D	18 Dec 2024 14:30	vial-A Internal Standard Fail	SY/MD	ReRun
14	P5299-04	SB-02	VY020635.D	18 Dec 2024 14:54	vial-A	SY/MD	Ok
15	P5301-03	HR-04-121624	VY020636.D	18 Dec 2024 15:17	vial-A Internal Standard Fail	SY/MD	ReRun
16	P5277-02	COMP-2A	VY020637.D	18 Dec 2024 15:41	vial-B NOT PURGE	SY/MD	Not Ok
17	P5301-01	HR-01-121624	VY020638.D	18 Dec 2024 16:04	vial-B	SY/MD	Ok
18	P5299-02	SB-02	VY020639.D	18 Dec 2024 16:27	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121824

Review By	Maresh Dadoda	Review On	12/19/2024 2:48:41 PM
Supervise By	Semsettin Yesilyurt	Supervise On	12/19/2024 2:49:14 PM
SubDirectory	VY121824	HP Acquire Method	HP Processing Method 82y121724s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132202		
CCC	VP132203,VP132204		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P5299-03	SB-01	VY020640.D	18 Dec 2024 16:51	vial-A	SY/MD	Ok
20	P5279-03	ROCKAWAY-PARK	VY020641.D	18 Dec 2024 17:14	vial-A NOT PURGE	SY/MD	Not Ok
21	P5299-01	SB-01	VY020642.D	18 Dec 2024 17:38	vial-A	SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY020643.D	18 Dec 2024 18:00		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121924

Review By	Maresh Dadoda	Review On	12/20/2024 10:52:52 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/20/2024 10:57:22 AM
SubDirectory	VY121924	HP Acquire Method	HP Processing Method 82y121724s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132238
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132239,VP132240 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020644.D	19 Dec 2024 09:13		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020645.D	19 Dec 2024 09:43		SY/MD	Ok,M
3	VY1219SBL01	VY1219SBL01	VY020646.D	19 Dec 2024 11:07		SY/MD	Ok
4	VY1219SBS01	VY1219SBS01	VY020647.D	19 Dec 2024 11:53	BS low for com #12	SY/MD	Ok,M
5	VY1219SBSD01	VY1219SBSD01	VY020648.D	19 Dec 2024 12:16		SY/MD	Ok,M
6	P5312-03RE	CONCRETE-VNJ-222RE	VY020649.D	19 Dec 2024 12:52	vial- B BS low; Internal standard fail	SY/MD	Confirms
7	P5312-01RE	SOIL-VNJ-222RE	VY020650.D	19 Dec 2024 13:15	vial- B BS low; Internal standard fail	SY/MD	Confirms
8	P5301-03RE	HR-04-121624RE	VY020651.D	19 Dec 2024 13:39	vial- B BS low; ISTD Fail	SY/MD	Confirms
9	P5306-01	OU4-VSL-07-121224	VY020652.D	19 Dec 2024 14:02	vial- B End CCC fail	SY/MD	ReRun
10	P5306-03	OU4-VSL-08-121224	VY020653.D	19 Dec 2024 14:26	vial- A End CCC fail; Not purge	SY/MD	Not Ok
11	P5306-05	OU4-VSL-09-121224	VY020654.D	19 Dec 2024 14:49	vial- A End CCC fail; Surrogate fail	SY/MD	ReRun
12	P5306-07	OU4-VSL-10-121224	VY020655.D	19 Dec 2024 15:12	vial- A End CCC fail; Internal standard fail; Surrogate fail	SY/MD	ReRun
13	P5306-09	OU4-VSL-11-121224	VY020656.D	19 Dec 2024 15:36	vial- A End CCC fail; Surrogate fail	SY/MD	ReRun
14	P5306-11	OU4-VSL-12-121224	VY020657.D	19 Dec 2024 15:59	vial- A End CCC fail; Surrogate fail	SY/MD	ReRun
15	P5306-13	OU4-VSL-13-121224	VY020658.D	19 Dec 2024 16:23	vial- A End CCC fail; Surrogate fail	SY/MD	ReRun
16	P5343-01	VNJ210	VY020659.D	19 Dec 2024 16:46	vial- A Internal standard fail	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121924

Review By	Maresh Dadoda	Review On	12/20/2024 10:52:52 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/20/2024 10:57:22 AM
SubDirectory	VY121924	HP Acquire Method	HP Processing Method 82y121724s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132238		
CCC	VP132239,VP132240		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	P5277-01	COMP-1A	VY020660.D	19 Dec 2024 17:09	vial- B	SY/MD	Ok
18	P5343-03	VNJ281	VY020661.D	19 Dec 2024 17:33		SY/MD	Ok
19	VSTDCCC050	VSTDCCC050EC	VY020662.D	19 Dec 2024 17:56	CCC fail for many comp.	SY/MD	Ok,M

M : Manual Integration