

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5277

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY020626.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1218SBS01	cis-1,2-Dichloroethene	20	19.2	ug/Kg	96			82	123	
	1,1,1-Trichloroethane	20	19.4	ug/Kg	97			80	126	
	Benzene	20	19.5	ug/Kg	98			84	121	
	Trichloroethene	20	19.4	ug/Kg	97			83	122	
	Toluene	20	19.3	ug/Kg	97			83	122	
	Ethyl Benzene	20	19.3	ug/Kg	97			82	124	
	m/p-Xylenes	40	38.3	ug/Kg	96			83	124	
	o-Xylene	20	19.1	ug/Kg	96			83	123	
	Isopropylbenzene	20	19.1	ug/Kg	96			82	124	