

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5277
Client: Kleinfelder
Analytical Method: SW8260D **Datafile :** VY020627.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1218SBSD01	cis-1,2-Dichloroethene	20	20.8	ug/Kg	104	8		82	123	20
	1,1,1-Trichloroethane	20	20.8	ug/Kg	104	7		80	126	20
	Benzene	20	20.7	ug/Kg	104	6		84	121	20
	Trichloroethene	20	20.7	ug/Kg	104	7		83	122	20
	Toluene	20	20.4	ug/Kg	102	5		83	122	20
	Ethyl Benzene	20	20.3	ug/Kg	102	5		82	124	20
	m/p-Xylenes	40	40.6	ug/Kg	102	6		83	124	20
	o-Xylene	20	20.6	ug/Kg	103	7		83	123	20
	Isopropylbenzene	20	20.1	ug/Kg	101	5		82	124	20