

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5277

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY020647.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1219SBS01	cis-1,2-Dichloroethene	20	21.3	ug/Kg	106			82	123	
	1,1,1-Trichloroethane	20	21.2	ug/Kg	106			80	126	
	Benzene	20	21.4	ug/Kg	107			84	121	
	Trichloroethene	20	21.4	ug/Kg	107			83	122	
	Toluene	20	21.5	ug/Kg	108			83	122	
	Ethyl Benzene	20	21.2	ug/Kg	106			82	124	
	m/p-Xylenes	40	42.9	ug/Kg	107			83	124	
	o-Xylene	20	21.7	ug/Kg	109			83	123	
	Isopropylbenzene	20	21.1	ug/Kg	106			82	124	