

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277
Lab File ID: VY020624.D Date Analyzed: 12/18/2024
Instrument ID: MSVOA_Y Time Analyzed: 10:06
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	168319	7.71	264221	8.62	226779	11.42
UPPER LIMIT	336638	8.213	528442	9.116	453558	11.92
LOWER LIMIT	84159.5	7.213	132111	8.116	113390	10.92
EPA SAMPLE NO.						
COMP-2A	167185	7.71	323384	8.62	272093	11.41
COMP-3A	165087	7.71	286183	8.62	239715	11.41
VY1218SBL01	201389	7.71	355898	8.62	296016	11.42
VY1218SBS01	167281	7.71	257615	8.62	226665	11.41
VY1218SBSD01	153098	7.71	244096	8.62	214944	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.