

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277
 Lab File ID: VY020645.D Date Analyzed: 12/19/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	206000	7.71	308825	8.62	266753	11.41
UPPER LIMIT	412000	8.207	617650	9.116	533506	11.914
LOWER LIMIT	103000	7.207	154413	8.116	133377	10.914
EPA SAMPLE NO.						
COMP-1A	108629	7.71	184812	8.62	172200	11.41
VY1219SBL01	204045	7.71	318952	8.62	271132	11.42
VY1219SBS01	184895	7.71	283194	8.62	247561	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.