

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: P5277 SAS No.: P5277 SDG NO.: P5277
 Lab File ID: VY020624.D Date Analyzed: 12/18/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:06
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	124266	13.347				
UPPER LIMIT	248532	13.847				
LOWER LIMIT	62133	12.847				
EPA SAMPLE NO.						
COMP-2A	105985	13.35				
COMP-3A	94502	13.35				
VY1218SBL01	114022	13.35				
VY1218SBS01	127451	13.35				
VY1218SBSD01	120747	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.