

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P5279
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5279 **SAS No.:** P5279
Matrix: Solid **Level:** LOW **Client ID:** ROCKAWAY-PARKA
Sample ID: P5279-01 **Spiked ID:** P5279-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	mg/Kg	75 - 125	68.8		0.34	J	37.7	182		P