



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc.

SDG No.: P5279

Contract: TULL02

Lab Code: CHEM

Case No.: P5279

SAS No.: P5279

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	85	40 - 160	CV	12/19/2024	12:07	LB134010
CRI01	Arsenic	21.1	20.0	106	40 - 160	P	12/20/2024	16:26	LB134054
	Barium	87.8	100	88	40 - 160	P	12/20/2024	16:26	LB134054
	Cadmium	5.88	6.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Chromium	10.2	10.0	102	40 - 160	P	12/20/2024	16:26	LB134054
	Lead	11.1	12.0	93	40 - 160	P	12/20/2024	16:26	LB134054
	Selenium	18.3	20.0	92	40 - 160	P	12/20/2024	16:26	LB134054
	Silver	10.3	10.0	103	40 - 160	P	12/20/2024	16:26	LB134054