

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P5279
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5279 **SAS No.:** P5279
Matrix: Water **Level:** LOW **Client ID:** ROCKAWAY-PARKA
Sample ID: P5279-02 **Spiked ID:** P5279-02A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	41.6		2.00	U	40.0	104		CV



Client:	<u>Tully Construction Co., Inc.</u>	SDG No.:	<u>P5279</u>				
Contract:	<u>TULL02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5279</u>	SAS No.:	<u>P5279</u>
Matrix:	<u>Water</u>	Level:	<u>LOW</u>	Client ID:	<u>WC-20241213A</u>		
Sample ID:	<u>P5291-13</u>	Spiked ID:	<u>P5291-13A</u>				

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Silver	ug/L	75 - 125	212		50.0	U	380	56		P