

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** TULL02

**Lab Code:** CHEM **Case No.:** P5279 **SAS No.:** P5279 **SDG NO.:** P5279

**Continuing Calib Date:** 12/16/2024 **Initial Calibration Date(s):** 12/06/2024 12/06/2024

**Continuing Calib Time:** 10:46 **Initial Calibration Time(s):** 14:19 22:34

**GC Column:** ZB-MR1 **ID:** 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.80       | 4.81      | 4.71      | 4.91 | 0.01       |
| Aroclor-1248-2 (2)   | 5.04       | 5.05      | 4.95      | 5.15 | 0.01       |
| Aroclor-1248-3 (3)   | 5.26       | 5.26      | 5.16      | 5.36 | 0.00       |
| Aroclor-1248-4 (4)   | 5.61       | 5.62      | 5.52      | 5.72 | 0.01       |
| Aroclor-1248-5 (5)   | 5.66       | 5.66      | 5.56      | 5.76 | 0.00       |
| Tetrachloro-m-xylene | 3.71       | 3.71      | 3.61      | 3.81 | 0.00       |
| Decachlorobiphenyl   | 8.77       | 8.79      | 8.69      | 8.89 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 10:46 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.79       | 4.80      | 4.70      | 4.90 | 0.01       |
| Aroclor-1248-2       | (2) | 5.03       | 5.03      | 4.93      | 5.13 | 0.00       |
| Aroclor-1248-3       | (3) | 5.07       | 5.08      | 4.98      | 5.18 | 0.01       |
| Aroclor-1248-4       | (4) | 5.24       | 5.25      | 5.15      | 5.35 | 0.01       |
| Aroclor-1248-5       | (5) | 5.63       | 5.64      | 5.54      | 5.74 | 0.01       |
| Tetrachloro-m-xylene |     | 3.70       | 3.71      | 3.61      | 3.81 | 0.01       |
| Decachlorobiphenyl   |     | 8.72       | 8.74      | 8.64      | 8.84 | 0.02       |