

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 21:30 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.80	4.81	4.71	4.91	0.01
Aroclor-1248-2 (2)	5.04	5.05	4.95	5.15	0.01
Aroclor-1248-3 (3)	5.26	5.26	5.16	5.36	0.00
Aroclor-1248-4 (4)	5.61	5.62	5.52	5.72	0.01
Aroclor-1248-5 (5)	5.66	5.66	5.56	5.76	0.00
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl	8.78	8.79	8.69	8.89	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5279 SAS No.: P5279 SDG NO.: P5279

Continuing Calib Date: 12/16/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 21:30 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1248-2	(2)	5.03	5.03	4.93	5.13	0.00
Aroclor-1248-3	(3)	5.07	5.08	4.98	5.18	0.01
Aroclor-1248-4	(4)	5.24	5.25	5.15	5.35	0.01
Aroclor-1248-5	(5)	5.64	5.64	5.54	5.74	0.00
Tetrachloro-m-xylene		3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl		8.73	8.74	8.64	8.84	0.01